


LB128726

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	-0.053	-0.053	333	0.000	333	0				6
0.2	0.200	0.212	0.012	1684	0.0 %	1684					6
2.5	2.500	2.644	0.144	14042	0.0 %	14042					6
5.0	5.000	4.944	-0.056	25731	0.0 %	25731					-1
7.5	7.500	7.366	-0.134	38043	0.0 %	38043					-2
10.0	10.000	10.087	0.087	51869	0.0 %	51869					1

LB128726 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Dec 2023 11:16:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	333	-	0.0000	7470A	PPB	14 Dec 2023 11:19:29	S	Std
0.2 - 1	50.2	1684	-	0.2000	7470A	PPB	14 Dec 2023 11:21:46	S	Std
2.5 - 1	52.5	14042	-	2.5000	7470A	PPB	14 Dec 2023 11:24:04	S	Std
5.0 - 1	55	25731	-	5.0000	7470A	PPB	14 Dec 2023 11:26:21	S	Std
7.5 - 1	57.5	38043	-	7.5000	7470A	PPB	14 Dec 2023 11:28:39	S	Std
10.0 - 1	510	51869	-	10.0000	7470A	PPB	14 Dec 2023 11:33:36	S	Std
ICV24 - 1	ICV24	21642	4.1393	-	7470A	PPB	14 Dec 2023 11:38:59	U	SMPL
ICB24 - 1	ICB24	240	-0.0717	-	7470A	PPB	14 Dec 2023 11:43:51	U	SMPL
CCV08 - 1	CCV08	24349	4.6720	-	7470A	PPB	14 Dec 2023 11:46:06	U	SMPL
CCB08 - 1	CCB08	122	-0.0949	-	7470A	PPB	14 Dec 2023 11:48:21	U	SMPL
CRA - 1	CRA	1632	0.2022	-	7470A	PPB	14 Dec 2023 11:50:38	U	SMPL
HighStd - 1	HighStd	51556	10.0252	-	7470A	PPB	14 Dec 2023 11:52:53	U	SMPL
ChkStd - 1	ChkStd	35363	6.8391	-	7470A	PPB	14 Dec 2023 11:55:08	U	SMPL
PB157766BL - 1	PBW	490	-0.0225	-	7470A	PPB	14 Dec 2023 12:00:03	U	SMPL
PB157766BS - 1	LCSW	21217	4.0557	-	7470A	PPB	14 Dec 2023 12:04:52	U	SMPL
O5758-02 - 1	SP-COMP-1	-274	-0.1729	-	7470A	PPB	14 Dec 2023 12:07:08	U	SMPL
O5758-11 - 1	SP-COMP-2	34	-0.1123	-	7470A	PPB	14 Dec 2023 12:09:27	U	SMPL
O5758-20 - 1	SP-COMP-3	38	-0.1115	-	7470A	PPB	14 Dec 2023 12:11:44	U	SMPL
O5758-29 - 1	SP-COMP-4	79	-0.1034	-	7470A	PPB	14 Dec 2023 12:14:01	U	SMPL
O5758-38 - 1	SP-COMP-5	87	-0.1018	-	7470A	PPB	14 Dec 2023 12:16:16	U	SMPL
CCV09 - 1	CCV09	24101	4.6232	-	7470A	PPB	14 Dec 2023 12:18:31	U	SMPL
CCB09 - 1	CCB09	266	-0.0666	-	7470A	PPB	14 Dec 2023 12:20:48	U	SMPL
O5767-02 - 1	SP-E1-COMP	42	-0.1107	-	7470A	PPB	14 Dec 2023 12:23:06	U	SMPL
O5767-05 - 1	SP-E2-COMP	76	-0.1040	-	7470A	PPB	14 Dec 2023 12:25:21	U	SMPL
O5767-08 - 1	SP-E3-COMP	85	-0.1022	-	7470A	PPB	14 Dec 2023 12:27:36	U	SMPL
O5768-01 - 1	ELZ-COMP	118	-0.0957	-	7470A	PPB	14 Dec 2023 12:29:52	U	SMPL
O5768-02 - 1	ELZ-23-00118	92	-0.1008	-	7470A	PPB	14 Dec 2023 12:32:07	U	SMPL
O5768-03 - 1	ELZ-23-00109	97	-0.0999	-	7470A	PPB	14 Dec 2023 12:34:24	U	SMPL
O5789-01 - 1	E8965C-5-001	198	-0.0800	-	7470A	PPB	14 Dec 2023 12:36:40	U	SMPL
O5791-04 - 1	597-P-2	100	-0.0993	-	7470A	PPB	14 Dec 2023 12:38:57	U	SMPL
O5791-09 - 1	MW-1B-SOIL	97	-0.0999	-	7470A	PPB	14 Dec 2023 12:41:14	U	SMPL
O5791-09DUP - 1	MW-1B-SOILDUP	115	-0.0963	-	7470A	PPB	14 Dec 2023 12:43:29	U	SMPL
CCV10 - 1	CCV10	25506	4.8996	-	7470A	PPB	14 Dec 2023 12:45:45	U	SMPL
CCB10 - 1	CCB10	328	-0.0544	-	7470A	PPB	14 Dec 2023 12:48:01	U	SMPL
O5791-09MS - 1	MW-1B-SOILMS	18505	3.5221	-	7470A	PPB	14 Dec 2023 12:50:19	U	SMPL
O5791-09MSD - 1	MW-1B-SOILMSD	17195	3.2643	-	7470A	PPB	14 Dec 2023 12:52:34	U	SMPL
PB157767BL - 1	PBW	249	-0.0700	-	7470A	PPB	14 Dec 2023 12:54:53	U	SMPL
PB157767BS - 1	LCSW	17126	3.2508	-	7470A	PPB	14 Dec 2023 12:57:11	U	SMPL
O5768-04 - 1	ELZ-23-00112	-165	-0.1514	-	7470A	PPB	14 Dec 2023 12:59:27	U	SMPL
O5768-04DUP - 1	ELZ-23-00112DUP	151	-0.0892	-	7470A	PPB	14 Dec 2023 13:01:46	U	SMPL
O5768-04MS - 1	ELZ-23-00112MS	19661	3.7496	-	7470A	PPB	14 Dec 2023 13:04:02	U	SMPL
O5768-04MSD - 1	ELZ-23-00112MSD	19896	3.7958	-	7470A	PPB	14 Dec 2023 13:06:19	U	SMPL
PB157730TB - 1	PB157730TB	378	-0.0446	-	7470A	PPB	14 Dec 2023 13:11:59	U	SMPL
PB157742TB - 1	PB157742TB	331	-0.0538	-	7470A	PPB	14 Dec 2023 13:14:15	U	SMPL
CCV11 - 1	CCV11	25272	4.8536	-	7470A	PPB	14 Dec 2023 13:16:31	U	SMPL
CCB11 - 1	CCB11	181	-0.0833	-	7470A	PPB	14 Dec 2023 13:18:46	U	SMPL
O5791-09LX5 - 1		-2	-0.1193	-	7470A	PPB	14 Dec 2023 13:21:05	U	SMPL
O5791-09A - 1		19185	3.6559	-	7470A	PPB	14 Dec 2023 13:23:21	U	SMPL
O5768-04LX5 - 1		-172	-0.1528	-	7470A	PPB	14 Dec 2023 13:25:36	U	SMPL
O5768-04A - 1		19656	3.7486	-	7470A	PPB	14 Dec 2023 13:27:54	U	SMPL
CCV12 - 1	CCV12	25412	4.8811	-	7470A	PPB	14 Dec 2023 13:30:10	U	SMPL
CCB12 - 1	CCB12	232	-0.0733	-	7470A	PPB	14 Dec 2023 13:32:28	U	SMPL