

LB128735


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

53492



A= 0.0000e+000

B= 1.8764e-004

C= -5.8746e-002

Rho= 0.9994286

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.020	-0.020	206	0.000	206	0				100
0.2	0.200	0.228	0.028	1528	0.0 %	1528					14
2.5	2.500	2.619	0.119	14268	0.0 %	14268					5
5.0	5.000	4.758	-0.242	25667	0.0 %	25667					2
7.5	7.500	7.637	0.137	41014	0.0 %	41014					0
10.0	10.000	9.979	-0.021	53492	0.0 %	53492					

LB128735 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Dec 2023 14:58:51

Sample ID	Extended ID	μ Abs	Conc	Strd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	206	-	0.000	7470A	PPB	14 Dec 2023 15:28:57	S	Std
0.2 - 1	50.2	1528	-	0.2000	7470A	PPB	14 Dec 2023 15:33:48	S	Std
2.5 - 1	52.5	14268	-	2.5000	7470A	PPB	14 Dec 2023 15:36:05	S	Std
5.0 - 1	55	25667	-	5.0000	7470A	PPB	14 Dec 2023 15:38:22	S	Std
7.5 - 1	57.5	41014	-	7.5000	7470A	PPB	14 Dec 2023 15:45:52	S	Std
10.0 - 1	58	53492	-	10.0000	7470A	PPB	14 Dec 2023 15:53:36	S	Std
ICV25 - 1	ICV25	20972	3.8765	-	7470A	PPB	14 Dec 2023 15:59:05	U	SMPL
ICB25 - 1	ICB25	-188	-0.0940	-	7470A	PPB	14 Dec 2023 16:01:21	U	SMPL
CCV13 - 1	CCV13	25131	4.6569	-	7470A	PPB	14 Dec 2023 16:03:39	U	SMPL
CCB13 - 1	CCB13	-183	-0.0931	-	7470A	PPB	14 Dec 2023 16:05:54	U	SMPL
CRA - 1	CRA	1505	0.2237	-	7470A	PPB	14 Dec 2023 16:08:11	U	SMPL
HighStd - 1	HighStd	55511	10.3576	-	7470A	PPB	14 Dec 2023 16:10:26	U	SMPL
ChkStd - 1	ChkStd	39284	7.3089	-	7470A	PPB	14 Dec 2023 16:12:42	U	SMPL
PB157803BL - 1	PBW	396	0.0156	-	7470A	PPB	14 Dec 2023 16:17:49	U	SMPL
PB157803BS - 1	LCSW	18991	3.5048	-	7470A	PPB	14 Dec 2023 16:20:05	U	SMPL
O4699-07 - 1	LOD-MDL-WATER-01-QT4-2023	533	0.0413	-	7470A	PPB	14 Dec 2023 16:33:36	U	SMPL
O4699-08 - 1	LOQ-WATER-02-QT4-2023	1558	0.2336	-	7470A	PPB	14 Dec 2023 16:35:52	U	SMPL
O5585-02 - 1	P3874-SP2	-38	-0.0659	-	7470A	PPB	14 Dec 2023 16:38:09	U	SMPL
O5585-03 - 1	P3874-SP3	118	-0.0366	-	7470A	PPB	14 Dec 2023 16:40:27	U	SMPL
O5791-01 - 1	597-A-J	404	0.0171	-	7470A	PPB	14 Dec 2023 16:42:42	U	SMPL
CCV14 - 1	CCV14	25703	4.7643	-	7470A	PPB	14 Dec 2023 16:44:57	U	SMPL
CCB14 - 1	CCB14	-118	-0.0809	-	7470A	PPB	14 Dec 2023 16:47:13	U	SMPL
O5791-01DUP - 1	597-A-JDUP	385	0.0135	-	7470A	PPB	14 Dec 2023 16:49:31	U	SMPL
O5791-01MS - 1	597-A-JMS	17589	3.2417	-	7470A	PPB	14 Dec 2023 16:51:47	U	SMPL
O5791-01MSD - 1	597-A-JMSD	18599	3.4313	-	7470A	PPB	14 Dec 2023 16:54:02	U	SMPL
O5791-02 - 1	597-K-O	-44	-0.0670	-	7470A	PPB	14 Dec 2023 16:56:20	U	SMPL
O5791-05 - 1	MW-1B-K-T	190	-0.0231	-	7470A	PPB	14 Dec 2023 16:58:39	U	SMPL
O5791-06 - 1	MW-1B-U-DD	302	-0.0021	-	7470A	PPB	14 Dec 2023 17:00:56	U	SMPL
O5791-07 - 1	MW-1B-00-XX	286	-0.0051	-	7470A	PPB	14 Dec 2023 17:03:12	U	SMPL
O5791-10 - 1	MW-1B-A-J	115	-0.0372	-	7470A	PPB	14 Dec 2023 17:05:30	U	SMPL
O5791-11 - 1	MW-1B-EE-NN	101	-0.0398	-	7470A	PPB	14 Dec 2023 17:07:46	U	SMPL
O5817-01 - 1	BOMARC-MW15-GW	143	-0.0319	-	7470A	PPB	14 Dec 2023 17:10:02	U	SMPL
CCV15 - 1	CCV15	25700	4.7637	-	7470A	PPB	14 Dec 2023 17:12:18	U	SMPL
CCB15 - 1	CCB15	-106	-0.0786	-	7470A	PPB	14 Dec 2023 17:14:34	U	SMPL
O5791-01LX5 - 1		-57	-0.0694	-	7470A	PPB	14 Dec 2023 17:16:52	U	SMPL
O5791-01A - 1		18418	3.3973	-	7470A	PPB	14 Dec 2023 17:19:08	U	SMPL
CCV16 - 1	CCV16	25510	4.7281	-	7470A	PPB	14 Dec 2023 17:21:23	U	SMPL
CCB16 - 1	CCB16	-139	-0.0848	-	7470A	PPB	14 Dec 2023 17:23:41	U	SMPL