

LB128743


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

70074



A= 0.0000e+000

B= 1.4610e-004

C= -6.2164e-002

Rho= 0.9991027

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1.1
0.0	0.000	-0.004	-0.004	398	0.000	398					16
0.2	0.200	0.231	0.031	2009	0.0 %	2009					8
2.5	2.500	2.690	0.190	18835	0.0 %	18835					4
5.0	5.000	4.776	-0.224	33112	0.0 %	33112					2
7.5	7.500	7.332	-0.168	50608	0.0 %	50608					2
10.0	10.000	10.176	0.176	70074	0.0 %	70074					

LB128743 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Dec 2023 17:45:11

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	398	-	0.0000	7471B	PPB	14 Dec 2023 17:49:15	S	Std
0.2 - 1	50.2	2009	-	0.2000	7471B	PPB	14 Dec 2023 17:53:53	S	Std
2.5 - 1	52.5	18835	-	2.5000	7471B	PPB	14 Dec 2023 17:56:11	S	Std
5.0 - 1	55	33112	-	5.0000	7471B	PPB	14 Dec 2023 17:58:29	S	Std
7.5 - 1	57.5	50608	-	7.5000	7471B	PPB	14 Dec 2023 18:00:51	S	Std
10.0 - 1	511	70074	-	10.0000	7471B	PPB	14 Dec 2023 18:11:21	S	Std
ICV26 - 1	ICV26	25921	3.7249	-	7471B	PPB	14 Dec 2023 18:14:15	U	SMPL
ICB26 - 1	ICB26	305	-0.0176	-	7471B	PPB	14 Dec 2023 18:16:31	U	SMPL
CCV17 - 1	CCV17	33646	4.8536	-	7471B	PPB	14 Dec 2023 18:18:56	U	SMPL
CCB17 - 1	CCB17	326	-0.0145	-	7471B	PPB	14 Dec 2023 18:21:10	U	SMPL
CRA - 1	CRA	2024	0.2335	-	7471B	PPB	14 Dec 2023 18:23:39	U	SMPL
HighStd - 1	HighStd	71021	10.3141	-	7471B	PPB	14 Dec 2023 18:25:54	U	SMPL
ChkStd - 1	ChkStd	46074	6.6893	-	7471B	PPB	14 Dec 2023 18:28:11	U	SMPL
PB157807BL - 1	PBS	-173	-0.0874	-	7471B	PPB	14 Dec 2023 18:30:43	U	SMPL
PB157807BS - 1	LCSS	26243	3.7720	-	7471B	PPB	14 Dec 2023 18:35:42	U	SMPL
O4699-01 - 1	LOD-MDL-SOIL-01-QT4-2023	707	0.0411	-	7471B	PPB	14 Dec 2023 18:40:39	U	SMPL
O4699-02 - 1	LOQ-SOIL-02-QT4-2023	1820	0.2037	-	7471B	PPB	14 Dec 2023 18:42:56	U	SMPL
O5759-03 - 1	AU-06-121323	7645	1.0548	-	7471B	PPB	14 Dec 2023 18:45:13	U	SMPL
O5790-01 - 1	72-11995	1831	0.2053	-	7471B	PPB	14 Dec 2023 18:47:29	U	SMPL
O5791-03 - 1	597-P-2	3985	0.5200	-	7471B	PPB	14 Dec 2023 18:49:47	U	SMPL
CCV18 - 1	CCV18	33407	4.8186	-	7471B	PPB	14 Dec 2023 18:52:03	U	SMPL
CCB18 - 1	CCB18	373	-0.0077	-	7471B	PPB	14 Dec 2023 18:54:21	U	SMPL
O5791-08 - 1	MW-1B-SOIL	1745	0.1928	-	7471B	PPB	14 Dec 2023 18:56:45	U	SMPL
O5793-01 - 1	OR-02-12122023	3738	0.4840	-	7471B	PPB	14 Dec 2023 18:59:00	U	SMPL
O5796-01 - 1	RT-SB-04-10-11	21011	3.0076	-	7471B	PPB	14 Dec 2023 19:01:16	U	SMPL
O5796-02 - 1	RT-SB-05-10.25-11.25	79971	11.6217	-	7471B	PPB	14 Dec 2023 19:03:34	U	SMPL
O5796-03 - 1	RT-SB-04-12-13	6872	0.9418	-	7471B	PPB	14 Dec 2023 19:05:55	U	SMPL
O5796-04 - 1	RT-SB-04-13-14	8065	1.1161	-	7471B	PPB	14 Dec 2023 19:08:28	U	SMPL
O5796-04DUP - 1	RT-SB-04-13-14DUP	7582	1.0458	-	7471B	PPB	14 Dec 2023 19:10:46	U	SMPL
O5796-05 - 1	O5796-04MS	36247	5.2338	-	7471B	PPB	14 Dec 2023 19:13:05	U	SMPL
O5796-06 - 1	O5796-04MSD	35695	5.1529	-	7471B	PPB	14 Dec 2023 19:17:46	U	SMPL
O5796-07 - 1	RT-SB-04-14-15	688	0.0384	-	7471B	PPB	14 Dec 2023 19:20:04	U	SMPL
CCV19 - 1	CCV19	34304	4.9497	-	7471B	PPB	14 Dec 2023 19:22:23	U	SMPL
CCB19 - 1	CCB19	532	0.0156	-	7471B	PPB	14 Dec 2023 19:24:39	U	SMPL
O5803-01 - 1	600572408-	2043	0.2363	-	7471B	PPB	14 Dec 2023 19:27:06	U	SMPL
O5805-01 - 1	HR-01-12122023	4265	0.5610	-	7471B	PPB	14 Dec 2023 19:29:21	U	SMPL
O5805-03 - 1	HR-04-12122023	2853	0.3547	-	7471B	PPB	14 Dec 2023 19:31:38	U	SMPL
O5796-04LX5 - 1		1741	0.1922	-	7471B	PPB	14 Dec 2023 19:33:56	U	SMPL
O5796-04A - 1		33349	4.8102	-	7471B	PPB	14 Dec 2023 19:36:13	U	SMPL
O5796-02DLX2 - 1		39809	5.7540	-	7471B	PPB	14 Dec 2023 19:38:29	U	SMPL
CCV20 - 1	CCV20	34692	5.0064	-	7471B	PPB	14 Dec 2023 19:40:57	U	SMPL
CCB20 - 1	CCB20	339	-0.0126	-	7471B	PPB	14 Dec 2023 19:43:19	U	SMPL