



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128749

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/15/2023	08:35
2	CAL2	1	Water	NA	NA	20.3	7.00	12/15/2023	08:36
3	CAL3	1	Water	NA	NA	20.3	10.02	12/15/2023	08:37
4	ICV	1	Water	NA	NA	20.5	7.01	12/15/2023	08:40
5	CCV1	1	Water	NA	NA	20.3	2.01	12/15/2023	08:44
6	O5784-03	1	Solid	20.02	20	23.3	5.43	12/15/2023	08:45
7	O5784-04	1	Solid	20.02	20	23.3	5.43	12/15/2023	08:45
8	O5784-05	1	Solid	20.02	20	23.3	5.43	12/15/2023	08:45
9	O5784-06	1	Solid	20.03	20	23.1	8.08	12/15/2023	08:55
10	O5784-07	1	Solid	20.04	20	22.7	7.85	12/15/2023	09:00
11	O5784-08	1	Solid	20.03	20	22.2	7.75	12/15/2023	09:05
12	O5784-09	1	Solid	20.04	20	23.0	7.61	12/15/2023	09:10
13	O5784-11	1	Solid	20.03	20	22.8	7.51	12/15/2023	09:15
14	O5784-12	1	Solid	20.04	20	22.6	7.55	12/15/2023	09:20
15	O5784-13	1	Solid	20.03	20	24.3	7.67	12/15/2023	09:22
16	CCV2	1	Water	NA	NA	20.3	12.02	12/15/2023	09:25
17	O5784-14	1	Solid	20.04	20	24.8	7.78	12/15/2023	09:30
18	O5784-16	1	Solid	20.02	20	23.2	7.72	12/15/2023	09:35
19	O5784-17	1	Solid	20.02	20	24.2	7.79	12/15/2023	09:40
20	O5784-20	1	Solid	20.02	20	23.5	7.73	12/15/2023	09:45

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O5784-21	1	Solid	20.03	20	23.4	8.19	12/15/2023	09:47
22	O5784-22	1	Solid	20.02	20	24.2	8.37	12/15/2023	09:50
23	CCV3	1	Water	NA	NA	20.3	2.01	12/15/2023	09:55