



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128750

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/15/2023	08:35
2	CAL2	1	Water	NA	NA	20.3	7.00	12/15/2023	08:36
3	CAL3	1	Water	NA	NA	20.3	10.02	12/15/2023	08:37
4	ICV	1	Water	NA	NA	20.3	7.00	12/15/2023	08:39
5	CCV1	1	Water	NA	NA	20.5	2.01	12/15/2023	08:40
6	O5831-03	1	Water	NA	NA	21.8	7.35	12/15/2023	08:45
7	O5831-03DUP	1	Water	NA	NA	21.8	7.36	12/15/2023	08:46
8	CCV2	1	Water	NA	NA	20.3	12.02	12/15/2023	08:49