



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128752

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.4	12/15/2023	08:20
2	O5768-04		1	12	100.00	>212.0	>212.0	12/15/2023	08:50
3	O5768-04DUP		1	12	100.00	>212.0	>212.0	12/15/2023	09:20
4	O5791-01		1	13	100.00	>212.0	>212.0	12/15/2023	09:50
5	O5791-02		1	14	100.00	>212.0	>212.0	12/15/2023	10:20
6	O5791-05		1	14	100.00	>212.0	>212.0	12/15/2023	10:50
7	O5791-06		1	12	100.00	>212.0	>212.0	12/15/2023	11:20
8	O5791-07		1	14	100.00	>212.0	>212.0	12/15/2023	11:50
9	O5791-10		1	12	100.00	>212.0	>212.0	12/15/2023	12:20
10	O5791-11		1	13	100.00	>212.0	>212.0	12/15/2023	12:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66128752

WorkList Name : fp-12-14 WorkList ID : 176353 Department : Wet-Chemistry Date : 12-14-2023 08:17:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5768-04	ELZ-23-00112	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/11/2023	1010B
O5791-01	597-A-J	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/12/2023	1010B
O5791-02	597-K-O	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/12/2023	1010B
O5791-05	MW-1B-K-T	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/12/2023	1010B
O5791-06	MW-1B-U-DD	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/12/2023	1010B
O5791-07	MW-1B-00-XX	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/12/2023	1010B
O5791-10	MW-1B-A-J	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/12/2023	1010B
O5791-11	MW-1B-EE-NN	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/12/2023	1010B

Date/Time 12/15/2023 08:15
 Raw Sample Received by: RM
 Raw Sample Relinquished by: RM

Date/Time 12/15/2023 13:00
 Raw Sample Received by: RM
 Raw Sample Relinquished by: RM