


LB124764

SFAM01.1

INSTRUMENT ID: CV4

Linear



A= 0.0000e+000

B= 1.6893e-004 slope

C= -1.5035e-001 y-intercept

Rho= 0.9985867

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	.0/0.0
0.00	0.000	-0.083	-0.083	400	0.000	400					
0.05	0.050					442	0				
0.20	0.200	0.187	-0.013	1995	0.0 %	1995					
2.50	2.500	2.414	-0.086	15179	0.0 %	15179					
5.00	5.000	5.243	0.243	31926	0.0 %	31926					
7.50	7.500	7.749	0.249	46759	0.0 %	46759					
10.0	10.000	9.691	-0.309	58257	0.0 %	58257					

LB128764

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	400	-		0 SFAM01.1	PPB	12/15/2023 16:21	Std
	0.2 S01	1995	-		0.2 SFAM01.1	PPB	12/15/2023 16:23	Std
	2.5 S02	15179	-		2.5 SFAM01.1	PPB	12/15/2023 16:28	Std
	5 S03	31926	-		5 SFAM01.1	PPB	12/15/2023 16:33	Std
	7.5 S04	46759	-		7.5 SFAM01.1	PPB	12/15/2023 16:38	Std
	10 S05	58257	-		10 SFAM01.1	PPB	12/15/2023 16:43	Std
ICV028	ICV028	23453	3.8115 -		SFAM01.1	PPB	12/15/2023 16:48	SMPL
ICB028	ICB028	583	-0.0519 -		SFAM01.1	PPB	12/15/2023 16:53	SMPL
CCV023	CCV023	31980	5.252 -		SFAM01.1	PPB	12/15/2023 16:55	SMPL
CCB023	CCB023	354	-0.0905 -		SFAM01.1	PPB	12/15/2023 16:58	SMPL
PB157850BL	PBW850	356	-0.0902 -		SFAM01.1	PPB	12/15/2023 17:00	SMPL
O5681-01	E0XP9	379	-0.0863 -		SFAM01.1	PPB	12/15/2023 17:02	SMPL
O5681-02	E0XQ9	367	-0.0884 -		SFAM01.1	PPB	12/15/2023 17:04	SMPL
O5681-03	E0XR0	476	-0.0699 -		SFAM01.1	PPB	12/15/2023 17:07	SMPL
O5681-04	E0XQ0	420	-0.0794 -		SFAM01.1	PPB	12/15/2023 17:09	SMPL
O5681-05	E0XQ1	826	-0.0108 -		SFAM01.1	PPB	12/15/2023 17:11	SMPL
O5681-06	E0XQ2	523	-0.062 -		SFAM01.1	PPB	12/15/2023 17:13	SMPL
O5681-07	E0XQ3	471	-0.0708 -		SFAM01.1	PPB	12/15/2023 17:16	SMPL
O5681-08	E0XR1	445	-0.0752 -		SFAM01.1	PPB	12/15/2023 17:18	SMPL
O5681-09	E0XR2	462	-0.0723 -		SFAM01.1	PPB	12/15/2023 17:20	SMPL
O5681-10	E0XR3	343	-0.0924 -		SFAM01.1	PPB	12/15/2023 17:23	SMPL
O5694-01	MJLAE1	359	-0.0897 -		SFAM01.1	PPB	12/15/2023 17:25	SMPL
CCV024	CCV024	32120	5.2756 -		SFAM01.1	PPB	12/15/2023 17:27	SMPL
CCB024	CCB024	391	-0.0843 -		SFAM01.1	PPB	12/15/2023 17:29	SMPL