



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128775

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/16/2023	11:35
2	CAL2	1	Water	NA	NA	20.3	7.00	12/16/2023	11:36
3	CAL3	1	Water	NA	NA	20.4	10.02	12/16/2023	11:37
4	ICV	1	Water	NA	NA	20.3	7.00	12/16/2023	11:40
5	CCV1	1	Water	NA	NA	20.3	2.01	12/16/2023	11:45
6	O5863-02	1	Water	NA	NA	21.6	9.46	12/16/2023	11:50
7	O5863-03	1	Water	NA	NA	24.1	7.57	12/16/2023	11:55
8	O5864-02	1	Water	NA	NA	20.5	7.98	12/16/2023	12:00
9	O5884-07	1	Water	NA	NA	23.4	8.62	12/16/2023	12:05
10	O5884-09	1	Water	NA	NA	23.4	6.55	12/16/2023	12:11
11	O5884-10	1	Water	NA	NA	23.6	11.45	12/16/2023	12:15
12	O5884-11	1	Water	NA	NA	23.6	8.20	12/16/2023	12:20
13	O5886-01	1	Water	NA	NA	23.2	5.13	12/16/2023	12:25
14	O5886-05	1	Water	NA	NA	23.6	8.80	12/16/2023	12:30
15	O5886-07	1	Water	NA	NA	23.2	4.87	12/16/2023	12:35
16	CCV2	1	Water	NA	NA	20.5	12.02	12/16/2023	12:40
17	O5886-09	1	Water	NA	NA	23.3	4.52	12/16/2023	12:45
18	O5886-09DUP	1	Water	NA	NA	23.4	4.53	12/16/2023	12:46
19	CCV3	1	Water	NA	NA	20.3	2.01	12/16/2023	12:50

WORKLIST(Hardcopy Internal Chain)

✓ 128775

WorkList Name : ph o5886

WorkList ID : 176450

Department : Wet-Chemistry

Date : 12-16-2023 11:21:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5863-02	D NWB-1987	Water	pH	Cool 4 deg C	PSEG03	N31	12/15/2023	9040C
O5863-03	C NWB-1932	Water	pH	Cool 4 deg C	PSEG03	N31	12/15/2023	9040C
O5864-02	A 1211-GW	Water	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9040C
O5884-07	A WATER-COMP	Water	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9040C
O5884-09	W-1	Water	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9040C
O5884-10	W-2	Water	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9040C
O5884-11	W-3	Water	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9040C
O5886-01	C AUD-1295	Water	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9040C
O5886-05	C AUD-1300	Water	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9040C
O5886-07	C AUD-1301	Water	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9040C
O5886-09	C AUD-1302	Water	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9040C

Date/Time 12-16-23 11:30

Raw Sample Received by: JDCSM

Raw Sample Relinquished by:

Date/Time 12-16-23

Raw Sample Received by: JDCSM

Raw Sample Relinquished by: JDCSM