



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128776

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.00	12/16/2023	11:30
2	CAL2	1	Water	NA	NA	20.3	7.00	12/16/2023	11:31
3	CAL3	1	Water	NA	NA	20.5	10.03	12/16/2023	11:33
4	ICV	1	Water	NA	NA	20.3	7.01	12/16/2023	11:35
5	CCV1	1	Water	NA	NA	20.5	2.01	12/16/2023	11:36
6	O5789-04	1	Solid	20.02	20	21.8	10.80	12/16/2023	11:40
7	O5803-03	1	Solid	20.03	20	21.7	6.91	12/16/2023	11:45
8	O5827-11	1	Solid	20.01	20	22.9	11.50	12/16/2023	11:47
9	O5853-03	1	Solid	20.02	20	21.0	7.68	12/16/2023	11:50
10	O5853-06	1	Solid	20.03	20	20.8	7.89	12/16/2023	11:55
11	O5862-11	1	Solid	20.02	20	22.2	9.27	12/16/2023	12:00
12	O5873-01	1	Solid	20.03	20	23.5	9.71	12/16/2023	12:10
13	O5887-04	1	Solid	20.02	20	21.0	6.21	12/16/2023	12:15
14	O5887-04DUP	1	Solid	20.03	20	21.1	6.22	12/16/2023	12:16
15	CCV2	1	Water	NA	NA	20.3	12.02	12/16/2023	12:20

VB 128716

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 05848

WorkList ID : 176448

Department: Wet-Chemistry

Date : 12-16-2023 11:19:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5789-04	E8965C-5-001	Solid	Corrosivity	Cool 4 deg C	ENTA05		12/12/2023	9045D
O5803-03	600572408-	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	12/13/2023	9045D
O5827-11	DS01-SO-121323	Solid	Corrosivity	Cool 4 deg C	LANG01	L11	12/13/2023	9045D
O5853-03	SOUTH-WALL	Solid	Corrosivity	Cool 4 deg C	TULL02	L21	12/14/2023	9045D
O5853-06	EAST-WALL	Solid	Corrosivity	Cool 4 deg C	TULL02	L21	12/14/2023	9045D
O5862-11	WC-SOLIDS-20231213-01	Solid	Corrosivity	Cool 4 deg C	PARS02	L21	12/13/2023	9045D
O5873-01	215-200	Solid	Corrosivity	Cool 4 deg C	WEST10	N11	12/11/2023	9045D
O5887-04	WC-121423	Solid	Corrosivity	Cool 4 deg C	PSEG03	N11	12/14/2023	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

12/16/23 11:20

30 (cc)

JDC SM

Date/Time 12/16/23

Raw Sample Received by:

Raw Sample Relinquished by:

141-00

JDC SM

30 (cc)