



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128777

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/16/2023	12:10
2	CAL2	1	Water	NA	NA	20.6	7.00	12/16/2023	12:11
3	CAL3	1	Water	NA	NA	20.3	10.02	12/16/2023	12:15
4	ICV	1	Water	NA	NA	20.3	7.00	12/16/2023	12:20
5	CCV1	1	Water	NA	NA	20.5	2.01	12/16/2023	12:25
6	O5862-13	1	Water	NA	NA	21.9	9.98	12/16/2023	12:35
7	O5872-01	1	Water	NA	NA	20.4	8.64	12/16/2023	12:40
8	O5872-01DUP	1	Water	NA	NA	20.5	8.65	12/16/2023	12:45
9	CCV2	1	Water	NA	NA	20.5	12.02	12/16/2023	12:50

VB128777

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corosovity O5838 water WorkList ID : 176449 Department : Wet-Chemistry Date : 12-16-2023 11:20:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5862-13	WC-LIQUIDS-20231213-01	Water	Corrosivity	Cool 4 deg C	PARS02	L21	12/13/2023	9040C
O5872-01	D 215-0207	Water	Corrosivity	Cool 4 deg C	WEST10	L31	12/13/2023	9040C

Date/Time 12.16.23 11:30
Raw Sample Received by: JDC (WC)
Raw Sample Relinquished by: JDC (SM)

Date/Time 12.16.23 16:00
Raw Sample Received by: JDC (SM)
Raw Sample Relinquished by: JDC (WC)