



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128778

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/16/2023	12:40
2	CAL2	1	Water	NA	NA	20.3	7.00	12/16/2023	12:41
3	CAL3	1	Water	NA	NA	20.3	10.02	12/16/2023	12:45
4	ICV	1	Water	NA	NA	20.5	7.02	12/16/2023	12:46
5	CCV1	1	Water	NA	NA	20.5	2.01	12/16/2023	12:50
6	O5756-02	1	Solid	20.02	20	25.4	6.70	12/16/2023	12:55
7	O5756-04	1	Solid	20.03	20	24.8	7.01	12/16/2023	12:56
8	O5756-06	1	Solid	20.04	20	24.2	7.56	12/16/2023	12:59
9	O5756-07	1	Solid	20.04	20	24.2	7.56	12/16/2023	12:59
10	O5756-08	1	Solid	20.04	20	24.2	7.56	12/16/2023	12:59
11	O5756-11	1	Solid	20.02	20	24.6	7.99	12/16/2023	13:11
12	O5756-12	1	Solid	20.03	20	24.4	8.01	12/16/2023	13:15
13	O5756-13	1	Solid	20.04	20	23.9	8.02	12/16/2023	13:20
14	O5756-14	1	Solid	20.02	20	23.8	8.02	12/16/2023	13:25
15	O5756-15	1	Solid	20.03	20	24.7	8.20	12/16/2023	13:30
16	CCV2	1	Water	NA	NA	20.5	12.02	12/16/2023	13:35
17	O5756-16	1	Solid	20.05	20	25.3	8.12	12/16/2023	13:40
18	O5756-17	1	Solid	20.02	20	24.5	8.54	12/16/2023	13:45
19	O5756-18	1	Solid	20.03	20	24.3	8.45	12/16/2023	13:50
20	O5756-20	1	Solid	20.02	20	25.2	8.45	12/16/2023	13:55

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	05756-21	1	Solid	20.03	20	24.6	8.27	12/16/2023	14:00
22	05756-22	1	Solid	20.04	20	24.8	8.44	12/16/2023	14:10
23	CCV3	1	Water	NA	NA	20.5	2.01	12/16/2023	14:15

WORKLIST(Hardcopy Internal Chain)

128778

WorkList Name : ph o5756

WorkList ID : 176452

Department : Wet-Chemistry

Date : 12-16-2023 12:26:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5756-02	E0ZA7	Solid	pH	Cool 4 deg C	USEP01	B33	12/06/2023	9045D
O5756-04	E0ZA9	Solid	pH	Cool 4 deg C	USEP01	B33	12/06/2023	9045D
O5756-06	E0ZQ2	Solid	pH	Cool 4 deg C	USEP01	B33	12/05/2023	9045D
O5756-07	E0ZQ2D	Solid	pH	Cool 4 deg C	USEP01	B33	12/05/2023	9045D
O5756-08	E0ZQ2S	Solid	pH	Cool 4 deg C	USEP01	B33	12/05/2023	9045D
O5756-11	E0YZ7	Solid	pH	Cool 4 deg C	USEP01	B33	12/04/2023	9045D
O5756-12	E0Z00	Solid	pH	Cool 4 deg C	USEP01	B33	12/04/2023	9045D
O5756-13	E0Z87	Solid	pH	Cool 4 deg C	USEP01	B33	12/06/2023	9045D
O5756-14	E0ZA6	Solid	pH	Cool 4 deg C	USEP01	B33	12/06/2023	9045D
O5756-15	E0ZA8	Solid	pH	Cool 4 deg C	USEP01	B33	12/06/2023	9045D
O5756-16	E0ZB0	Solid	pH	Cool 4 deg C	USEP01	B33	12/06/2023	9045D
O5756-17	E0ZF9	Solid	pH	Cool 4 deg C	USEP01	B33	12/07/2023	9045D
O5756-18	E0ZG0	Solid	pH	Cool 4 deg C	USEP01	B33	12/07/2023	9045D
O5756-20	E0Z61	Solid	pH	Cool 4 deg C	USEP01	B33	12/05/2023	9045D
O5756-21	E0Z62	Solid	pH	Cool 4 deg C	USEP01	B33	12/05/2023	9045D
O5756-22	E0Z65	Solid	pH	Cool 4 deg C	USEP01	B33	12/05/2023	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: