



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 12/18/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 12/17/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 12/18/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB128784

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5780-01	EZQA2	1	1.18	8.81	9.99	8.2	79.7	
O5780-02	EZQA4	2	1.19	8.62	9.81	8.29	82.4	
O5780-03	EZQA5	3	1.18	8.46	9.64	8.12	82.0	
O5780-04	EZQA6	4	1.15	8.51	9.66	7.8	78.1	
O5780-05	EZQA7	5	1.16	8.40	9.56	7.96	81.0	
O5780-06	EZQA8	6	1.15	8.83	9.98	8.92	88.0	
O5780-07	EZQA9	7	1.15	8.78	9.93	7.88	76.7	
O5780-08	EZQB0	8	1.19	8.59	9.78	7.86	77.6	
O5780-09	EZQB1	9	1.19	8.63	9.82	7.86	77.3	
O5780-10	EZQB2	10	1.17	8.60	9.77	8.2	81.7	
O5780-11	EZQB3	11	1.19	8.43	9.62	8.11	82.1	
O5780-12	EZQB4	12	1.19	8.58	9.77	8.27	82.5	
O5780-13	EZQB5	13	1.14	8.82	9.96	8.7	85.7	
O5780-14	EZQB6	14	1.18	8.48	9.66	8.12	81.8	
O5780-15	EZQB7	15	1.18	8.50	9.68	8.2	82.6	
O5780-16	EZQB8	16	1.17	8.67	9.84	8.34	82.7	
O5780-17	EZQB9	17	1.15	8.43	9.58	7.93	80.4	
O5780-18	EZQC0	18	1.15	8.52	9.67	8.15	82.2	
O5780-19	EZQC1	19	1.15	8.37	9.52	8.18	84.0	
O5780-20	EZQC6	20	1.16	8.53	9.69	8.27	83.4	
O5780-21	EZQC6D	21	1.16	8.53	9.69	8.27	83.4	
O5780-22	EZQC6S	22	1.16	8.53	9.69	8.27	83.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$