



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128792

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	12/17/2023	08:25
2	O5831-03		1	12	100.00	>212.0	>212.0	12/17/2023	08:50
3	O5862-13		1	14	100.00	>212.0	>212.0	12/17/2023	09:20
4	O5863-02		1	12	100.00	>212.0	>212.0	12/17/2023	09:50
5	O5863-03		1	12	100.00	>212.0	>212.0	12/17/2023	10:20
6	O5863-04		1	4	8.00	46.4	45.7	12/17/2023	10:50
7	O5863-04DUP		1	4	8.00	46.4	45.7	12/17/2023	11:20
8	O5864-02		1	14	100.00	>212.0	>212.0	12/17/2023	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16128792

WorkList Name : fp-12-17

WorkList ID : 176463

Department : Wet-Chemistry

Date : 12-17-2023 08:02:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5831-03	A6931	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/14/2023	1010B
O5862-13 	WC-LIQUIDS-20231213-01	Water	Flash Point	Cool 4 deg C	PARS02	L21	12/13/2023	1010B
O5863-02 	NWB-1987	Water	Flash Point	Cool 4 deg C	PSEG03	N31	12/15/2023	1010B
O5863-03 	NWB-1932	Water	Flash Point	Cool 4 deg C	PSEG03	N31	12/15/2023	1010B
O5863-04 	NWB-19534	Water	Flash Point	Cool 4 deg C	PSEG03	N31	12/15/2023	1010B
O5864-02 	1211-GW	Water	Flash Point	Cool 4 deg C	PSEG03	N11	12/15/2023	1010B

Date/Time 12/17/2023 08:15
 Raw Sample Received by: RM gwc
 Raw Sample Relinquished by: M. A. E.

Date/Time 12/17/2023 12:30
 Raw Sample Received by: J. A. C. 1
 Raw Sample Relinquished by: RM cwc