

LB128803


7471B

INSTRUMENT ID: CV

Linear

 μ Abs.:

65739



A= 0.0000e+000

B= 1.5278e-004

C= -4.7226e-002

Rho= 0.9999773

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	G/D
0.0	0.000	0.008	0.008	362	0.000	362					1
0.2	0.200	0.222	0.022	1760	0.0 %	1760					1
2.5	2.500	2.447	-0.053	16326	0.0 %	16326					2
5.0	5.000	5.016	0.016	33142	0.0 %	33142					0
7.5	7.500	7.510	0.010	49464	0.0 %	49464					0
10.0	10.000	9.997	-0.003	65739	0.0 %	65739					0

LB128803 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Dec 2023 12:29:58

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	362	-	0.0000	7471B	PPB	18 Dec 2023 13:10:52	S	Std
0.2 - 1	50-2	1760	-	0.2000	7471B	PPB	18 Dec 2023 13:13:09	S	Std
2.5 - 1	52.5	16326	-	2.5000	7471B	PPB	18 Dec 2023 13:15:26	S	Std
5.0 - 1	55	33142	-	5.0000	7471B	PPB	18 Dec 2023 13:17:42	S	Std
7.5 - 1	57.5	49464	-	7.5000	7471B	PPB	18 Dec 2023 13:20:02	S	Std
10.0 - 1	511	65739	-	10.0000	7471B	PPB	18 Dec 2023 13:22:28	S	Std
ICV30 - 1	ICV30	26800	4.0474	-	7471B	PPB	18 Dec 2023 13:25:30	U	SMPL
ICB30 - 1	ICB30	-157	-0.0712	-	7471B	PPB	18 Dec 2023 13:27:47	U	SMPL
CCV27 - 1	CCV27	33563	5.0807	-	7471B	PPB	18 Dec 2023 13:30:10	U	SMPL
CCB27 - 1	CCB27	-162	-0.0720	-	7471B	PPB	18 Dec 2023 13:32:25	U	SMPL
CRA - 1	CRA	1665	0.2072	-	7471B	PPB	18 Dec 2023 13:34:50	U	SMPL
HighStd - 1	HighStd	66663	10.1378	-	7471B	PPB	18 Dec 2023 13:37:06	U	SMPL
ChkStd - 1	ChkStd	46091	6.9948	-	7471B	PPB	18 Dec 2023 13:39:22	U	SMPL
PB157857BL - 1	PBS	-304	-0.0937	-	7471B	PPB	18 Dec 2023 13:41:52	U	SMPL
PB157857BS - 1	LCSS	29964	4.5308	-	7471B	PPB	18 Dec 2023 13:44:19	U	SMPL
O5798-01 - 1	WC-1	4299	0.6096	-	7471B	PPB	18 Dec 2023 13:46:35	U	SMPL
O5827-01 - 1	RT-DUP01-121223	1124	0.1245	-	7471B	PPB	18 Dec 2023 13:48:59	U	SMPL
O5827-02 - 1	RT-SB-07-14-15	865	0.0849	-	7471B	PPB	18 Dec 2023 13:51:18	U	SMPL
O5827-03 - 1	RT-SB-07-10-11	22621	3.4089	-	7471B	PPB	18 Dec 2023 13:53:34	U	SMPL
O5827-04 - 1	RT-SB-07-23.5-24.5	304	-0.0008	-	7471B	PPB	18 Dec 2023 13:55:50	U	SMPL
CCV0128 - 1	CCV0128	34168	5.1731	-	7471B	PPB	18 Dec 2023 13:58:11	U	SMPL
CCB0128 - 1	CCB0128	50	-0.0396	-	7471B	PPB	18 Dec 2023 14:00:27	U	SMPL
O5827-05 - 1	RT-SB-08-13.5-14.5	3579	0.4996	-	7471B	PPB	18 Dec 2023 14:02:51	U	SMPL
O5827-06 - 1	RT-SB-08-12.5-13.5	5231	0.7520	-	7471B	PPB	18 Dec 2023 14:05:07	U	SMPL
O5827-07 - 1	RT-SB-09-13.75-14.75	772	0.0707	-	7471B	PPB	18 Dec 2023 14:07:24	U	SMPL
O5827-07DUP - 1	RT-SB-09-13.75-14.75DUP	863	0.0846	-	7471B	PPB	18 Dec 2023 14:09:42	U	SMPL
O5827-07MS - 1	RT-SB-09-13.75-14.75MS	30305	4.5829	-	7471B	PPB	18 Dec 2023 14:11:58	U	SMPL
O5827-07MSD - 1	RT-SB-09-13.75-14.75MSD	30219	4.5698	-	7471B	PPB	18 Dec 2023 14:14:14	U	SMPL
O5827-08 - 1	RT-SB-08-15.5-16.5	8376	1.2325	-	7471B	PPB	18 Dec 2023 14:16:37	U	SMPL
O5827-09 - 1	RT-SB-09-10.5-11.5	102917	15.6769	-	7471B	PPB	18 Dec 2023 14:19:01	U	SMPL
O5827-10 - 1	RT-SB-10-11-12	2754	0.3735	-	7471B	PPB	18 Dec 2023 14:21:20	U	SMPL
O5831-01 - 1	CHRT-20017	579477	HIGH	-	7471B	PPB	18 Dec 2023 14:23:52	X	
CCV29 - 1	CCV29	34890	5.2834	-	7471B	PPB	18 Dec 2023 14:25:39	U	SMPL
CCB29 - 1	CCB29	-261	-0.0871	-	7471B	PPB	18 Dec 2023 14:28:19	U	SMPL
O5832-01 - 1	CHRT-20103	3422	0.4756	-	7471B	PPB	18 Dec 2023 14:30:42	U	SMPL
O5833-01 - 1	351	5543	0.7997	-	7471B	PPB	18 Dec 2023 14:32:58	U	SMPL
O5853-01 - 1	SOUTH-WALL	64805	9.8540	-	7471B	PPB	18 Dec 2023 14:35:16	U	SMPL
O5853-04 - 1	EAST-WALL	94642	14.4126	-	7471B	PPB	18 Dec 2023 14:37:33	U	SMPL
O5862-10 - 1	WC-SOLIDS-20231213-01	3485	0.4852	-	7471B	PPB	18 Dec 2023 14:40:01	U	SMPL
O5827-07LX5 - 1		23	-0.0437	-	7471B	PPB	18 Dec 2023 14:42:32	U	SMPL
O5827-07A - 1		29270	4.4248	-	7471B	PPB	18 Dec 2023 14:44:52	U	SMPL
O5827-09DLX5 - 1		23263	3.5070	-	7471B	PPB	18 Dec 2023 14:47:08	U	SMPL
O5831-01DLX10 - 1		97576	14.8609	-	7471B	PPB	18 Dec 2023 14:49:32	U	SMPL
O5853-04DLX5 - 1		20612	3.1020	-	7471B	PPB	18 Dec 2023 14:51:50	U	SMPL
CCV30 - 1	CCV30	34772	5.2654	-	7471B	PPB	18 Dec 2023 14:54:22	U	SMPL
CCB30 - 1	CCB00	97	-0.0324	-	7471B	PPB	18 Dec 2023 14:56:42	U	SMPL
O6831-01DL2X50 - 1		16482	2.4710	-	7471B	PPB	18 Dec 2023 14:59:06	U	SMPL
CCV31 - 1	CCV31	33953	5.1403	-	7471B	PPB	18 Dec 2023 15:01:22	U	SMPL
CCB31 - 1	CCB31	145	-0.0251	-	7471B	PPB	18 Dec 2023 15:03:40	U	SMPL