



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 12/19/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:05
In Date: 12/18/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: m oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 12/19/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: m sc-4
Thermometer ID: % solids-oven

QC:LB128807

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5856-01	MBGS80	1	1.13	8.66	9.79	8.28	82.6	
O5856-02	MBGS81	2	1.18	8.58	9.76	8.21	81.9	
O5856-03	MBGS82	3	1.15	8.57	9.72	8.48	85.5	
O5856-04	MBGSA5	4	1.18	8.55	9.73	8.06	80.5	
O5856-05	MBGSA6	5	1.14	8.65	9.79	7.99	79.2	
O5856-06	MBGRD1	6	1.18	8.79	9.97	8.87	87.5	
O5856-07	MBGSA7	7	1.12	8.46	9.58	8.03	81.7	
O5856-08	MBGSA8	8	1.19	8.43	9.62	7.99	80.7	
O5856-09	MBGSC5	9	1.13	8.49	9.62	8.3	84.5	
O5856-10	MBGSC6	10	1.17	8.60	9.77	8.73	87.9	
O5856-11	MBGSC7	11	1.19	8.63	9.82	8.86	88.9	
O5856-12	MBGSC8	12	1.15	8.72	9.87	8.87	88.5	
O5856-13	MBGSC9	13	1.15	8.82	9.97	8.65	85.0	
O5856-14	MBGSD0	14	1.18	8.64	9.82	8.38	83.3	
O5856-15	MBGSD1	15	1.16	8.47	9.63	8.02	81.0	
O5856-16	MBGSD2	16	1.15	8.40	9.55	8.13	83.1	
O5856-17	MBGRD2	17	1.12	8.57	9.69	8.66	88.0	
O5856-18	MBGRW2	18	1.17	8.74	9.91	9.19	91.8	
O5856-19	MBGRW3	19	1.13	8.61	9.74	8.94	90.7	
O5856-20	MBGRX1	20	1.13	8.43	9.56	8.79	90.9	
O5856-21	MBGRX1D	21	1.13	8.43	9.56	8.79	90.9	
O5856-22	MBGRX1S	22	1.13	8.43	9.56	8.79	90.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$