



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 12/19/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 12/18/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 12/19/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB128809

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5859-01	BGRD0	1	1.15	8.42	9.57	8.57	88.1	
O5859-02	BGRX9	2	1.18	8.48	9.66	8.35	84.6	
O5859-03	BGRY0	3	1.18	8.73	9.91	8.96	89.1	
O5859-04	BGS95	4	1.14	8.77	9.91	8.54	84.4	
O5859-05	BGS96	5	1.19	8.58	9.77	8.35	83.4	
O5859-06	BGS97	6	1.18	8.42	9.6	8.12	82.4	
O5859-07	BGS98	7	1.18	8.50	9.68	8.03	80.6	
O5859-08	BGS99	8	1.18	8.37	9.55	8.35	85.7	
O5859-09	BGSA0	9	1.14	8.40	9.54	8.77	90.8	
O5859-10	BGSA1	10	1.18	8.79	9.97	9.00	89.0	
O5859-11	BGRY1	11	1.18	8.77	9.95	8.8	86.9	
O5859-12	BGRY2	12	1.14	8.63	9.77	8.82	89.0	
O5859-13	BGRY3	13	1.16	8.62	9.78	9.11	92.2	
O5859-14	BGRY4	14	1.19	8.44	9.63	9.05	93.1	
O5859-15	BGRY5	15	1.15	8.40	9.55	8.9	92.3	
O5859-16	BGRY6	16	1.18	8.47	9.65	8.96	91.9	
O5859-17	BGRY7	17	1.13	8.56	9.69	8.38	84.7	
O5859-18	BGRY7MS	18	1.13	8.56	9.69	8.38	84.7	
O5859-19	BGRY7MSD	19	1.13	8.56	9.69	8.38	84.7	
O5859-20	BGRY8	20	1.16	8.50	9.66	8.73	89.1	
O5859-21	BGSA3	21	1.18	8.37	9.55	8.4	86.3	
O5859-22	BGSA4	22	1.19	8.43	9.62	8.73	89.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$