



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 12/19/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 12/18/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: m oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 12/19/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: m sc-4
Thermometer ID: % solids-oven

QC:LB128810

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5854-01	MBGRD0	1	1.15	8.42	9.57	8.57	88.1	
O5854-02	MBGRX9	2	1.18	8.48	9.66	8.35	84.6	
O5854-03	MBGRY0	3	1.18	8.73	9.91	8.96	89.1	
O5854-04	MBGS95	4	1.14	8.77	9.91	8.54	84.4	
O5854-05	MBGS96	5	1.19	8.58	9.77	8.35	83.4	
O5854-06	MBGS97	6	1.18	8.42	9.6	8.12	82.4	
O5854-07	MBGS98	7	1.18	8.50	9.68	8.03	80.6	
O5854-08	MBGS99	8	1.18	8.37	9.55	8.35	85.7	
O5854-09	MBGSA0	9	1.14	8.40	9.54	8.77	90.8	
O5854-10	MBGSA1	10	1.18	8.79	9.97	9.00	89.0	
O5854-11	MBGRY1	11	1.18	8.77	9.95	8.8	86.9	
O5854-12	MBGRY2	12	1.14	8.63	9.77	8.82	89.0	
O5854-13	MBGRY3	13	1.16	8.62	9.78	9.11	92.2	
O5854-14	MBGRY4	14	1.19	8.44	9.63	9.05	93.1	
O5854-15	MBGRY5	15	1.15	8.40	9.55	8.9	92.3	
O5854-16	MBGRY6	16	1.18	8.47	9.65	8.96	91.9	
O5854-17	MBGRY7	17	1.13	8.56	9.69	8.38	84.7	
O5854-18	MBGRY7D	18	1.13	8.56	9.69	8.38	84.7	
O5854-19	MBGRY7S	19	1.13	8.56	9.69	8.38	84.7	
O5854-20	MBGRY8	20	1.16	8.50	9.66	8.73	89.1	
O5854-21	MBGSA3	21	1.18	8.37	9.55	8.4	86.3	
O5854-22	MBGSA4	22	1.19	8.43	9.62	8.73	89.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$