



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 12/19/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 13:15  
In Date: 12/18/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:50  
Out Date: 12/19/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB128811

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O5849-01 | E0664           | 1      | 1.13           | 8.83         | 9.96                    | 8.81                      | 87.0    |          |
| O5849-02 | E0665           | 2      | 1.15           | 8.38         | 9.53                    | 8.35                      | 85.9    |          |
| O5849-03 | E0666           | 3      | 1.18           | 8.57         | 9.75                    | 7.53                      | 74.1    |          |
| O5849-04 | E0667           | 4      | 1.12           | 8.39         | 9.51                    | 7.96                      | 81.5    |          |
| O5849-05 | E0668           | 5      | 1.17           | 8.35         | 9.52                    | 7.96                      | 81.3    |          |
| O5849-06 | E0669           | 6      | 1.18           | 8.53         | 9.71                    | 8.33                      | 83.8    |          |
| O5849-07 | E0670           | 7      | 1.14           | 8.42         | 9.56                    | 7.43                      | 74.7    |          |
| O5849-08 | E0671           | 8      | 1.12           | 8.56         | 9.68                    | 8.17                      | 82.4    |          |
| O5849-09 | E0672           | 9      | 1.16           | 8.80         | 9.96                    | 8.48                      | 83.2    |          |
| O5849-10 | E0673           | 10     | 1.15           | 8.38         | 9.53                    | 8.16                      | 83.7    |          |
| O5849-11 | E0674           | 11     | 1.14           | 8.78         | 9.92                    | 8.56                      | 84.5    |          |
| O5849-12 | E0675           | 12     | 1.15           | 8.80         | 9.95                    | 8.17                      | 79.8    |          |
| O5849-13 | E0676           | 13     | 1.17           | 8.60         | 9.77                    | 8.26                      | 82.4    |          |
| O5849-14 | E0677           | 14     | 1.15           | 8.41         | 9.56                    | 7.99                      | 81.3    |          |
| O5849-15 | E0677MS         | 15     | 1.15           | 8.41         | 9.56                    | 7.99                      | 81.3    |          |
| O5849-16 | E0677MSD        | 16     | 1.15           | 8.41         | 9.56                    | 7.99                      | 81.3    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$