



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128814

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/18/2023	15:15
2	CAL2	1	Water	NA	NA	20.3	7.00	12/18/2023	15:16
3	CAL3	1	Water	NA	NA	20.3	10.02	12/18/2023	15:17
4	ICV	1	Water	NA	NA	20.4	7.00	12/18/2023	15:20
5	CCV1	1	Water	NA	NA	20.3	2.01	12/18/2023	15:25
6	O5886-03	1	Solid	20.02	20	23.2	6.03	12/18/2023	15:35
7	O5886-03DUP	1	Solid	20.01	20	23.3	6.03	12/18/2023	15:36
8	CCV2	1	Water	NA	NA	20.5	12.02	12/18/2023	15:40

128814

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5886-1

WorkList ID : 176508

Department : Wet-Chemistry

Date : 12-18-2023 15:04:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5886-03	AUD-1299	Solid	pH	Cool 4 deg C	PSEG03	N11	12/15/2023	9045D

12-18-23 15:10

Date/Time

Raw Sample Received by: JH CWC

Raw Sample Relinquished by: JH CWC

Date/Time 12-18-23

Raw Sample Received by: JH CWC

Raw Sample Relinquished by: JH CWC