



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128817

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/18/2023	15:20
2	CAL2	1	Water	NA	NA	20.5	7.00	12/18/2023	15:21
3	CAL3	1	Water	NA	NA	20.3	10.02	12/18/2023	15:25
4	ICV	1	Water	NA	NA	20.5	7.02	12/18/2023	15:30
5	CCV1	1	Water	NA	NA	20.5	2.01	12/18/2023	15:35
6	O5910-01	1	Solid	20.02	20	24.4	8.19	12/18/2023	15:40
7	O5910-01DUP	1	Solid	20.03	20	24.5	8.20	12/18/2023	15:41
8	CCV2	1	Water	NA	NA	20.4	12.02	12/18/2023	15:45

WORKLIST(Hardcopy Internal Chain)

✓ 126817

WorkList Name : CORROSIVITY O5910

WorkList ID : 176507

Department : Wet-Chemistry

Date : 12-18-2023 15:03:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5910-01	OILY-STONE	Solid	Corrosivity	Cool 4 deg C	PSEG03	O21	12/18/2023	9045D

Date/Time 12-18-23 15:10

Raw Sample Received by: JAC

Raw Sample Relinquished by: JAC

Date/Time 12-18-23 17:00

Raw Sample Received by: JAC

Raw Sample Relinquished by: JAC