



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128819

Ambient Barometric Pressure (mmHg): 740.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.6	12/18/2023	08:30
2	O5872-01		1	12	100.00	>212.0	>212.0	12/18/2023	09:00
3	O5884-07		1	14	100.00	>212.0	>212.0	12/18/2023	09:30
4	O5886-01		1	14	100.00	>212.0	>212.0	12/18/2023	10:00
5	O5886-05		1	12	100.00	>212.0	>212.0	12/18/2023	10:30
6	O5886-07		1	12	100.00	>212.0	>212.0	12/18/2023	11:00
7	O5886-09		1	12	100.00	>212.0	>212.0	12/18/2023	11:30
8	O5886-09DUP		1	12	100.00	>212.0	>212.0	12/18/2023	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6128819

WorkList Name : fp 1218

WorkList ID : 176491

Department : Wet-Chemistry

Date : 12-18-2023 08:10:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5872-01	215-0207	Water	Flash Point	Cool 4 deg C	WEST10	L31	12/13/2023	1010B
O5884-07	WATER-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N11	12/15/2023	1010B
O5886-01	AUD-1295	Water	Flash Point	Cool 4 deg C	PSEG03	N11	12/15/2023	1010B
O5886-05	AUD-1300	Water	Flash Point	Cool 4 deg C	PSEG03	N11	12/15/2023	1010B
O5886-07	AUD-1301	Water	Flash Point	Cool 4 deg C	PSEG03	N11	12/15/2023	1010B
O5886-09	AUD-1302	Water	Flash Point	Cool 4 deg C	PSEG03	N11	12/15/2023	1010B

Date/Time 12/18/2023 08:15
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JDC CWC

Date/Time 12/18/2023 12:40
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JDC CWC