

LB128823


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

61009



A= 0.0000e+000

B= 1.6296e-004

C= 3.5077e-002

Rho= 0.9998013

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.056	0.056	131	0.000	131					17
0.2	0.200	0.233	0.033	1217	0.0 %	1217					-4
2.5	2.500	2.399	-0.101	14505	0.0 %	14505					-1
5.0	5.000	4.926	-0.074	30011	0.0 %	30011					1
7.5	7.500	7.609	0.109	46477	0.0 %	46477					0
10.0	10.000	9.977	-0.023	61009	0.0 %	61009					

LB128823 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Dec 2023 08:29:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	131	-	0.0000	7471B	PPB	19 Dec 2023 08:36:19	S	Std
0.2 - 1	50.2	1217	-	0.2000	7471B	PPB	19 Dec 2023 08:38:36	S	Std
2.5 - 1	52.5	14505	-	2.5000	7471B	PPB	19 Dec 2023 08:40:53	S	Std
5.0 - 1	55	30011	-	5.0000	7471B	PPB	19 Dec 2023 08:43:10	S	Std
7.5 - 1	57.5	46477	-	7.5000	7471B	PPB	19 Dec 2023 08:45:29	S	Std
10.0 - 1	510	61009	-	10.0000	7471B	PPB	19 Dec 2023 08:47:55	S	Std
ICV31 - 1	ICV31	24197	3.9782	-	7471B	PPB	19 Dec 2023 08:50:57	U	SMPL
ICB31 - 1	ICB31	-262	-0.0076	-	7471B	PPB	19 Dec 2023 08:53:13	U	SMPL
CCV32 - 1	CCV32	30842	5.0610	-	7471B	PPB	19 Dec 2023 08:55:35	U	SMPL
CCB32 - 1	CCB32	-256	-0.0066	-	7471B	PPB	19 Dec 2023 08:57:50	U	SMPL
CRA - 1	CRA	1165	0.2249	-	7471B	PPB	19 Dec 2023 09:00:15	U	SMPL
HighStd - 1	HighStd	61041	9.9822	-	7471B	PPB	19 Dec 2023 09:02:30	U	SMPL
ChkStd - 1	ChkStd	44300	7.2541	-	7471B	PPB	19 Dec 2023 09:04:45	U	SMPL
PB157890BL - 1	PBS	-122	0.0152	-	7471B	PPB	19 Dec 2023 09:07:16	U	SMPL
PB157890BS - 1	LCSS	27765	4.5596	-	7471B	PPB	19 Dec 2023 09:09:45	U	SMPL
O5882-01 - 1	TR-04-121523	3568	0.6165	-	7471B	PPB	19 Dec 2023 09:12:01	U	SMPL
O5883-01 - 1	EO-01-121523	3707	0.6392	-	7471B	PPB	19 Dec 2023 09:14:27	U	SMPL
O5883-03 - 1	EO-02-121523	4854	0.8261	-	7471B	PPB	19 Dec 2023 09:16:47	U	SMPL
O5884-01 - 1	SOIL-COMP-1	3947	0.6783	-	7471B	PPB	19 Dec 2023 09:19:06	U	SMPL
O5884-03 - 1	SOIL-COMP-2	674	0.1449	-	7471B	PPB	19 Dec 2023 09:21:23	U	SMPL
CCV33 - 1	CCV33	30947	5.0781	-	7471B	PPB	19 Dec 2023 09:23:41	U	SMPL
CCB33 - 1	CCB33	-111	0.0170	-	7471B	PPB	19 Dec 2023 09:25:56	U	SMPL
O5884-05 - 1	SOIL-COMP-3	999	0.1979	-	7471B	PPB	19 Dec 2023 09:28:22	U	SMPL
O5887-01 - 1	WC-121423	6358	1.0712	-	7471B	PPB	19 Dec 2023 09:30:37	U	SMPL
O5887-01DUP - 1	WC-121423DUP	5740	0.9705	-	7471B	PPB	19 Dec 2023 09:32:53	U	SMPL
O5887-01MS - 1	WC-121423MS	35573	5.8320	-	7471B	PPB	19 Dec 2023 09:35:10	U	SMPL
O5887-01MSD - 1	WC-121423MSD	32616	5.3501	-	7471B	PPB	19 Dec 2023 09:37:28	U	SMPL
O5887-01LX5 - 1		1096	0.2137	-	7471B	PPB	19 Dec 2023 09:39:55	U	SMPL
O5887-01A - 1		30214	4.9587	-	7471B	PPB	19 Dec 2023 09:42:22	U	SMPL
CCV34 - 1	CCV34	31361	5.1456	-	7471B	PPB	19 Dec 2023 09:44:39	U	SMPL
CCB34 - 1	CCB34	-59	0.0255	-	7471B	PPB	19 Dec 2023 09:47:04	U	SMPL