


LB128825

7470A

INSTRUMENT ID: CW



Linear

A= 0.0000e+000

B= 2.1901e-004

C= 1.3659e-002

Rho= 0.9996028

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.038	0.038	113	0.000	113					4/00
0.2	0.200	0.191	-0.009	811	0.0 %	811					-5
2.5	2.500	2.342	-0.158	10629	0.0 %	10629					-6
5.0	5.000	5.180	0.180	23589	0.0 %	23589					4
7.5	7.500	7.522	0.022	34284	0.0 %	34284					0
10.0	10.000	9.927	-0.073	45263	0.0 %	45263					-1

LB128825 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Dec 2023 10:36:41

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	113	-	0.0000	7470A	PPB	19 Dec 2023 10:41:45	S	Std
0.2 - 1	50.2	811	-	0.2000	7470A	PPB	19 Dec 2023 10:44:03	S	Std
2.5 - 1	52.5	10629	-	2.5000	7470A	PPB	19 Dec 2023 10:46:20	S	Std
5.0 - 1	55	23589	-	5.0000	7470A	PPB	19 Dec 2023 10:48:37	S	Std
7.5 - 1	57.5	34284	-	7.5000	7470A	PPB	19 Dec 2023 10:50:56	S	Std
10.0 - 1	59.5	45263	-	10.0000	7470A	PPB	19 Dec 2023 10:53:15	S	Std
ICV32 - 1	ICV32	16497	3.6267	-	7470A	PPB	19 Dec 2023 10:58:01	U	SMPL
ICB32 - 1	ICB32	-256	-0.0424	-	7470A	PPB	19 Dec 2023 10:58:17	U	SMPL
CCV35 - 1	CCV35	21694	4.7649	-	7470A	PPB	19 Dec 2023 11:00:35	U	SMPL
CCB35 - 1	CCB35	-253	-0.0418	-	7470A	PPB	19 Dec 2023 11:02:52	U	SMPL
CRA - 1	CRA	695	0.1659	-	7470A	PPB	19 Dec 2023 11:05:09	U	SMPL
HighStd - 1	HighStd	45064	9.8831	-	7470A	PPB	19 Dec 2023 11:07:24	U	SMPL
ChkStd - 1	ChkStd	29691	6.5163	-	7470A	PPB	19 Dec 2023 11:09:40	U	SMPL
PB157880BL - 1	PBW	-502	-0.0963	-	7470A	PPB	19 Dec 2023 11:11:58	U	SMPL
PB157880BS - 1	LCSW	17448	3.8349	-	7470A	PPB	19 Dec 2023 11:14:18	U	SMPL
O5827-11 - 1	DS01-SO-121323	-324	-0.0573	-	7470A	PPB	19 Dec 2023 11:16:34	U	SMPL
O5853-03 - 1	SOUTH-WALL	-58	0.0010	-	7470A	PPB	19 Dec 2023 11:18:53	U	SMPL
O5853-06 - 1	EAST-WALL	-51	0.0025	-	7470A	PPB	19 Dec 2023 11:21:10	U	SMPL
O5862-11 - 1	WC-SOLIDS-20231213-01	-26	0.0080	-	7470A	PPB	19 Dec 2023 11:23:28	U	SMPL
O5873-01 - 1	215-200	0	0.0137	-	7470A	PPB	19 Dec 2023 11:25:43	U	SMPL
CCV36 - 1	CCV36	22253	4.8873	-	7470A	PPB	19 Dec 2023 11:27:58	U	SMPL
CCB36 - 1	CCB36	-231	-0.0369	-	7470A	PPB	19 Dec 2023 11:30:13	U	SMPL
O5881-02 - 1	TRE-1564	-81	-0.0041	-	7470A	PPB	19 Dec 2023 11:32:31	U	SMPL
O5881-04 - 1	TRE-1565	-46	0.0036	-	7470A	PPB	19 Dec 2023 11:34:48	U	SMPL
O5886-04 - 1	AUD-1299	1662	0.3777	-	7470A	PPB	19 Dec 2023 11:37:02	U	SMPL
O5887-04 - 1	WC-121423	-149	-0.0190	-	7470A	PPB	19 Dec 2023 11:39:18	U	SMPL
O5887-04DUP - 1	WC-121423DUP	-6	0.0123	-	7470A	PPB	19 Dec 2023 11:41:37	U	SMPL
O5887-04MS - 1	WC-121423MS	18408	4.0452	-	7470A	PPB	19 Dec 2023 11:43:53	U	SMPL
O5887-04MSD - 1	WC-121423MSD	17714	3.8932	-	7470A	PPB	19 Dec 2023 11:46:10	U	SMPL
PB157881BL - 1	PBW	-279	-0.0474	-	7470A	PPB	19 Dec 2023 11:48:29	U	SMPL
PB157881BS - 1	LCSW	17686	3.8871	-	7470A	PPB	19 Dec 2023 11:50:48	U	SMPL
O5862-13 - 1	WC-LIQUIDS-20231213-01	-274	-0.0463	-	7470A	PPB	19 Dec 2023 11:53:04	U	SMPL
CCV37 - 1	CCV37	24136	5.2997	-	7470A	PPB	19 Dec 2023 11:55:21	U	SMPL
CCB37 - 1	CCB37	-334	-0.0595	-	7470A	PPB	19 Dec 2023 11:57:37	U	SMPL
O5862-13DUP - 1	WC-LIQUIDS-20231213-01DUP	-56	0.0014	-	7470A	PPB	19 Dec 2023 11:59:55	U	SMPL
O5862-13MS - 1	WC-LIQUIDS-20231213-01MS	18026	3.9615	-	7470A	PPB	19 Dec 2023 12:02:12	U	SMPL
O5862-13MSD - 1	WC-LIQUIDS-20231213-01MSD	18198	3.9992	-	7470A	PPB	19 Dec 2023 12:04:27	U	SMPL
O5872-01 - 1	215-0207	-380	-0.0696	-	7470A	PPB	19 Dec 2023 12:06:46	U	SMPL
O5886-02 - 1	AUD-1295	1715	0.3893	-	7470A	PPB	19 Dec 2023 12:09:04	U	SMPL
O5886-06 - 1	AUD-1300	-27	0.0077	-	7470A	PPB	19 Dec 2023 12:11:21	U	SMPL
O5886-08 - 1	AUD-1301	644	0.1547	-	7470A	PPB	19 Dec 2023 12:13:38	U	SMPL
O5886-10 - 1	AUD-1302	53	0.0253	-	7470A	PPB	19 Dec 2023 12:15:55	U	SMPL
PB157882BL - 1	PBW	202	0.0579	-	7470A	PPB	19 Dec 2023 12:18:12	U	SMPL
PB157882BS - 1	LCSW	19621	4.3109	-	7470A	PPB	19 Dec 2023 12:20:28	U	SMPL
CCV38 - 1	CCV38	24834	5.4526	-	7470A	PPB	19 Dec 2023 12:36:07	U	SMPL
CCB38 - 1	CCB38	-273	-0.0461	-	7470A	PPB	19 Dec 2023 12:38:22	U	SMPL
O5831-03 - 1	A6931	9725	2.1435	-	7470A	PPB	19 Dec 2023 12:40:40	U	SMPL
O5831-03DUP - 1	A6931DUP	9497	2.0936	-	7470A	PPB	19 Dec 2023 12:42:56	U	SMPL
O5831-03MS - 1	A6931MS	23833	5.2333	-	7470A	PPB	19 Dec 2023 12:45:13	U	SMPL
O5831-03MSD - 1	A6931MSD	23838	5.2344	-	7470A	PPB	19 Dec 2023 12:47:31	U	SMPL
O5862-12 - 1	WC-LIQUIDS-20231213-01	2167	0.4883	-	7470A	PPB	19 Dec 2023 12:49:49	U	SMPL
O5864-02 - 1	1211-GW	-57	0.0012	-	7470A	PPB	19 Dec 2023 12:52:08	U	SMPL
O5884-07 - 1	WATER-COMP	55	0.0257	-	7470A	PPB	19 Dec 2023 12:54:26	U	SMPL
PB157841TB - 1	PB157841TB	-25	0.0082	-	7470A	PPB	19 Dec 2023 12:56:43	U	SMPL
PB157858TB - 1	PB157858TB	7	0.0152	-	7470A	PPB	19 Dec 2023 12:59:00	U	SMPL
CCV39 - 1	CCV39	22826	5.0128	-	7470A	PPB	19 Dec 2023 13:01:16	U	SMPL
CCB39 - 1	CCB39	-208	-0.0319	-	7470A	PPB	19 Dec 2023 13:03:31	U	SMPL
O5887-04LX5 - 1		-54	0.0018	-	7470A	PPB	19 Dec 2023 13:05:49	U	SMPL
O5887-04A - 1		18629	4.0936	-	7470A	PPB	19 Dec 2023 13:08:05	U	SMPL
O5862-13LX5 - 1		-227	-0.0361	-	7470A	PPB	19 Dec 2023 13:10:20	U	SMPL
O5862-13A - 1		18431	4.0502	-	7470A	PPB	19 Dec 2023 13:12:38	U	SMPL
O5831-03LX5 - 1		2103	0.4742	-	7470A	PPB	19 Dec 2023 13:14:54	U	SMPL
O5831-03A - 1		27232	5.9777	-	7470A	PPB	19 Dec 2023 13:17:12	U	SMPL
CCV40 - 1	CCV40	24502	5.3798	-	7470A	PPB	19 Dec 2023 13:19:29	U	SMPL
CCB40 - 1	CCB40	-302	-0.0525	-	7470A	PPB	19 Dec 2023 13:21:48	U	SMPL