



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128831

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/19/2023	12:45
2	CAL2	1	Water	NA	NA	20.3	7.00	12/19/2023	12:46
3	CAL3	1	Water	NA	NA	20.3	10.02	12/19/2023	12:48
4	ICV	1	Water	NA	NA	20.4	7.02	12/19/2023	12:50
5	CCV1	1	Water	NA	NA	20.5	2.01	12/19/2023	12:55
6	O5839-01	1	Solid	20.02	20	24.4	10.50	12/19/2023	13:00
7	O5839-02	1	Solid	20.01	20	23.6	10.98	12/19/2023	13:10
8	O5839-03	1	Solid	20.01	20	23.6	10.98	12/19/2023	13:10
9	O5839-04	1	Solid	20.01	20	23.6	10.98	12/19/2023	13:10
10	O5839-05	1	Solid	20.02	20	23.3	11.18	12/19/2023	13:15
11	O5839-06	1	Solid	20.02	20	23.2	10.88	12/19/2023	13:20
12	O5839-07	1	Solid	20.03	20	23.4	10.25	12/19/2023	13:25
13	O5839-08	1	Solid	20.03	20	22.6	11.15	12/19/2023	13:30
14	O5839-09	1	Solid	20.02	20	22.6	11.15	12/19/2023	13:35
15	O5839-10	1	Solid	20.02	20	23.2	10.90	12/19/2023	13:36
16	CCV2	1	Water	NA	NA	20.3	12.02	12/19/2023	13:38
17	O5839-11	1	Solid	20.03	20	23.2	10.99	12/19/2023	13:40
18	O5839-12	1	Solid	20.02	20	22.1	10.44	12/19/2023	13:45
19	O5839-15	1	Solid	20.03	20	22.9	9.42	12/19/2023	13:47
20	O5839-16	1	Solid	20.02	20	22.9	12.34	12/19/2023	13:50

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	05839-17	1	Solid	20.01	20	23.5	9.68	12/19/2023	14:00
22	05839-18	1	Solid	20.02	20	22.7	12.02	12/19/2023	14:05
23	05839-19	1	Solid	20.01	20	23.6	9.35	12/19/2023	14:10
24	05839-20	1	Solid	20.02	20	23.4	10.60	12/19/2023	14:15
25	05839-21	1	Solid	20.05	20	23.3	10.97	12/19/2023	14:20
26	05839-22	1	Solid	20.03	20	24.4	9.41	12/19/2023	14:15
27	CCV3	1	Water	NA	NA	20.5	2.01	12/19/2023	14:20

Analytical Summary Report

Analysis Method: 9045D
 Parameter: pH
 Run Number: LB128831
 BalanceID: WC SC-4

Analyst By : jignesh
 Supervisor Review By : Iwona
 Slope : 99.2
 pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER, COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/19/2023	12:45
2	CAL2	1	Water	NA	NA	20.3	7.00	12/19/2023	12:46
3	CAL3	1	Water	NA	NA	20.3	10.02	12/19/2023	12:48
4	ICV	1	Water	NA	NA	20.4	7.02	12/19/2023	12:50
5	CCV1	1	Water	NA	NA	20.5	2.01	12/19/2023	12:55
6	O5839-01	1	Solid	20.02	20	24.4	10.50	12/19/2023	13:00
7	O5839-02	1	Solid	20.01	20	23.6	10.98	12/19/2023	13:10
8	O5839-03	1	Solid	20.01	20	23.6	10.98	12/19/2023	13:10
9	O5839-04	1	Solid	20.01	20	23.6	10.98	12/19/2023	13:10
10	O5839-05	1	Solid	20.02	20	23.3	11.18	12/19/2023	13:15
11	O5839-06	1	Solid	20.02	20	23.2	10.88	12/19/2023	13:20
12	O5839-07	1	Solid	20.03	20	23.4	10.25	12/19/2023	13:25
13	O5839-08	1	Solid	20.03	20	22.6	11.15	12/19/2023	13:30
14	O5839-09	1	Solid	20.02	20	22.6	11.15	12/19/2023	13:35
15	O5839-10	1	Solid	20.02	20	23.2	10.90	12/19/2023	13:36
16	CCV2	1	Water	NA	NA	20.3	12.02	12/19/2023	13:38
17	O5839-11	1	Solid	20.03	20	23.2	10.99	12/19/2023	13:40
18	O5839-12	1	Solid	20.02	20	22.1	10.44	12/19/2023	13:45
19	O5839-15	1	Solid	20.03	20	22.9	9.42	12/19/2023	13:47
20	O5839-16	1	Solid	20.02	20	22.9	12.34	12/19/2023	13:50

20
 01/04/24
 25

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O5839-17	1	Solid	20.01	20	23.5	9.68	12/19/2023	14:00
22	O5839-18	1	Solid	20.02	20	22.7	12.02	12/19/2023	14:05
23	O5839-19	1	Solid	20.01	20	23.6	9.35	12/19/2023	14:10
24	O5839-20	1	Solid	20.02	20	23.4	10.60	12/19/2023	14:15
25	O5839-21	1	Solid	20.05	20	23.3	10.97	12/19/2023	14:20
26	O5839-22	1	Solid	20.03	20	24.4	9.41	12/19/2023	14:15
27	CCV3	1	Water	NA	NA	20.5	2.01	12/19/2023	14:20