


L3128846

SFAM01.1

INSTRUMENT ID: EV1



Linear

A= 0.0000e+000

B= 1.8925e-004 slope

C= -7.7092e-003 + intercept

Rho= 0.9998004

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.025	0.025	175	0.000	175					
0.05	0.050					442	0				
0.20	0.200	0.206	0.006	1131	0.0 %	1131					
2.50	2.500	2.488	-0.012	13190	0.0 %	13190					
5.00	5.000	4.893	-0.107	25898	0.0 %	25898					
7.50	7.500	7.635	0.135	40383	0.0 %	40383					
10.0	10.000	9.952	-0.048	52628	0.0 %	52628					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		175	-		0 SFAM01.1	PPB	12/20/2023 9:12	Std
0.2		1131	-		0.2 SFAM01.1	PPB	12/20/2023 9:14	Std
2.5		13190	-		2.5 SFAM01.1	PPB	12/20/2023 9:17	Std
5		25898	-		5 SFAM01.1	PPB	12/20/2023 9:19	Std
7.5		40383	-		7.5 SFAM01.1	PPB	12/20/2023 9:21	Std
10		52628	-		10 SFAM01.1	PPB	12/20/2023 9:37	Std
ICV033	ICV033	20607	3.8921 -		SFAM01.1	PPB	12/20/2023 9:40	SMPL
ICB033	ICB033	-153	-0.0367 -		SFAM01.1	PPB	12/20/2023 9:42	SMPL
CCV040	CCV040	27325	5.1634 -		SFAM01.1	PPB	12/20/2023 9:44	SMPL
CCB040	CCB040	-167	-0.0393 -		SFAM01.1	PPB	12/20/2023 9:46	SMPL
PB157935BL	PBW935	116	0.0142 -		SFAM01.1	PPB	12/20/2023 9:51	SMPL
O5753-01	EOZN0	1599	0.2949 -		SFAM01.1	PPB	12/20/2023 9:54	SMPL
O5753-02	EOZN1	1517	0.2794 -		SFAM01.1	PPB	12/20/2023 9:56	SMPL
O5753-03	EOZP0	1489	0.2741 -		SFAM01.1	PPB	12/20/2023 9:58	SMPL
O5753-04	EOZP1	411	0.0701 -		SFAM01.1	PPB	12/20/2023 10:00	SMPL
O5753-05	EOZM7	1672	0.3087 -		SFAM01.1	PPB	12/20/2023 10:03	SMPL
O5753-06	EOZN7	1566	0.2886 -		SFAM01.1	PPB	12/20/2023 10:05	SMPL
O5753-11	EOZM8	1465	0.2695 -		SFAM01.1	PPB	12/20/2023 10:07	SMPL
O5753-12	EOZM9	1387	0.2548 -		SFAM01.1	PPB	12/20/2023 10:09	SMPL
O5753-13	EOZN8	1921	0.3558 -		SFAM01.1	PPB	12/20/2023 10:12	SMPL
O5753-14	EOZN9	423	0.0723 -		SFAM01.1	PPB	12/20/2023 10:14	SMPL
PB157936BL	PBW936	404	0.0687 -		SFAM01.1	PPB	12/20/2023 10:16	SMPL
O5868-01	GCF06	435	0.0746 -		SFAM01.1	PPB	12/20/2023 10:19	SMPL
O5868-02	GCF06D	419	0.0716 -		SFAM01.1	PPB	12/20/2023 10:21	SMPL
O5868-03	GCF06S	5981	1.1242 -		SFAM01.1	PPB	12/20/2023 10:23	SMPL
O5868-04	GCF07	1139	0.2078 -		SFAM01.1	PPB	12/20/2023 10:25	SMPL
O5868-05	GCF08	458	0.079 -		SFAM01.1	PPB	12/20/2023 10:28	SMPL
O5868-06	GCF09	407	0.0693 -		SFAM01.1	PPB	12/20/2023 10:30	SMPL
O5868-07	GCF10	306	0.0502 -		SFAM01.1	PPB	12/20/2023 10:32	SMPL
O5868-08	GCF11	442	0.0759 -		SFAM01.1	PPB	12/20/2023 10:34	SMPL
CCV041	CCV041	26831	5.0699 -		SFAM01.1	PPB	12/20/2023 10:37	SMPL
CCB041	CCB041	-106	-0.0278 -		SFAM01.1	PPB	12/20/2023 10:39	SMPL
O5868-09	GCF12	108	0.0127 -		SFAM01.1	PPB	12/20/2023 10:41	SMPL
O5868-10	GCF13	148	0.0203 -		SFAM01.1	PPB	12/20/2023 10:44	SMPL
O5868-11	GCF14	164	0.0233 -		SFAM01.1	PPB	12/20/2023 10:46	SMPL
O5868-12	GCF15	183	0.0269 -		SFAM01.1	PPB	12/20/2023 10:48	SMPL
CCV042	CCV042	25526	4.823 -		SFAM01.1	PPB	12/20/2023 11:16	SMPL
CCB042	CCB042	-88	-0.0244 -		SFAM01.1	PPB	12/20/2023 11:18	SMPL