



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128855

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	12/20/2023	11:15
2	O5863-06		1	4	8.00	46.4	45.4	12/20/2023	11:45
3	O5863-06DUP		1	4	8.00	46.4	45.4	12/20/2023	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6128855

WorkList Name : fp-s-1220

WorkList ID : 176577

Department : Wet-Chemistry

Date : 12-20-2023 09:56:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5863-06	NWB-19534	Solid	Flash Point	Cool 4 deg C	PSEG03	N31	12/15/2023	1010B

Date/Time 12/20/2023 11:10
 Raw Sample Received by: RM curc
 Raw Sample Relinquished by: 7p/curc

Date/Time 12/20/2023 12:50
 Raw Sample Received by: 7p/curc
 Raw Sample Relinquished by: RM curc