


LB128857

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.6283e-004 *slope*

C= -7.1970e-002 *y-intercept*

Rho= 0.9996248

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>total</i>
0.00	0.000	-0.025	-0.025	290	0.000	290					
0.05	0.050					442	0				-16
0.20	0.200	0.169	-0.031	1479	0.0 %	1479					-1
2.50	2.500	2.487	-0.013	15714	0.0 %	15714					1
5.00	5.000	5.055	0.055	31488	0.0 %	31488					2
7.50	7.500	7.675	0.175	47576	0.0 %	47576					-2
10.0	10.000	9.839	-0.161	60867	0.0 %	60867					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	290	-		0 SFAM01.1	PPB	12/20/2023 13:58	Std
	0.2 S01	1479	-		0.2 SFAM01.1	PPB	12/20/2023 14:00	Std
	2.5 S02	15714	-		2.5 SFAM01.1	PPB	12/20/2023 14:02	Std
	5 S03	31488	-		5 SFAM01.1	PPB	12/20/2023 14:05	Std
	7.5 S04	47576	-		7.5 SFAM01.1	PPB	12/20/2023 14:07	Std
	10 S05	60867	-		10 SFAM01.1	PPB	12/20/2023 14:11	Std
ICV034	ICV034	23154	3.6982 -		SFAM01.1	PPB	12/20/2023 14:14	SMPL
ICB034	ICB034	5	-0.0712 -		SFAM01.1	PPB	12/20/2023 14:16	SMPL
CCV043	CCV043	30764	4.9374 -		SFAM01.1	PPB	12/20/2023 14:18	SMPL
CCB043	CCB043	-40	-0.0785 -		SFAM01.1	PPB	12/20/2023 14:21	SMPL
PB157932BL	PBS932	104	-0.055 -		SFAM01.1	PPB	12/20/2023 14:23	SMPL
O5721-02	EOZ13	3509	0.4994 -		SFAM01.1	PPB	12/20/2023 14:25	SMPL
O5721-03	EOZJ6	454	0.002 -		SFAM01.1	PPB	12/20/2023 14:27	SMPL
O5721-04	EOZJ7	1856	0.2302 -		SFAM01.1	PPB	12/20/2023 14:30	SMPL
O5721-05	EOZJ8	2160	0.2797 -		SFAM01.1	PPB	12/20/2023 14:32	SMPL
O5721-06	EOZJ9	2047	0.2613 -		SFAM01.1	PPB	12/20/2023 14:34	SMPL
O5721-07	EOZK0	1349	0.1477 -		SFAM01.1	PPB	12/20/2023 14:36	SMPL
O5721-08	EOZP7	1275	0.1356 -		SFAM01.1	PPB	12/20/2023 14:39	SMPL
O5721-11	EOZ41	2716	0.3703 -		SFAM01.1	PPB	12/20/2023 14:41	SMPL
O5721-12	EOZ42	1409	0.1575 -		SFAM01.1	PPB	12/20/2023 14:43	SMPL
O5721-13	EOZ43	438	-0.0006 -		SFAM01.1	PPB	12/20/2023 14:46	SMPL
O5721-14	EOZ44	385	-0.0093 -		SFAM01.1	PPB	12/20/2023 14:48	SMPL
O5721-15	EOZ45	656	0.0348 -		SFAM01.1	PPB	12/20/2023 14:50	SMPL
O5721-16	EOZ37	1667	0.1995 -		SFAM01.1	PPB	12/20/2023 14:52	SMPL
O5721-17	EOZ40	1153	0.1158 -		SFAM01.1	PPB	12/20/2023 14:55	SMPL
O5721-18	EOZ56	2003	0.2542 -		SFAM01.1	PPB	12/20/2023 14:57	SMPL
O5721-19	EOZ57	1078	0.1036 -		SFAM01.1	PPB	12/20/2023 14:59	SMPL
O5721-20	EOZP9	2700	0.3677 -		SFAM01.1	PPB	12/20/2023 15:01	SMPL
O5721-21	EOZP9D	2859	0.3936 -		SFAM01.1	PPB	12/20/2023 15:04	SMPL
O5721-22	EOZP9S	18830	2.9941 -		SFAM01.1	PPB	12/20/2023 15:06	SMPL
CCV044	CCV044	30103	4.8297 -		SFAM01.1	PPB	12/20/2023 15:08	SMPL
CCB044	CCB044	30	-0.0671 -		SFAM01.1	PPB	12/20/2023 15:11	SMPL
PB157933BL	PBS933	87	-0.0578 -		SFAM01.1	PPB	12/20/2023 15:13	SMPL
O5722-01	EOZ58	933	0.08 -		SFAM01.1	PPB	12/20/2023 15:15	SMPL
O5722-02	EOZ59	1523	0.176 -		SFAM01.1	PPB	12/20/2023 15:18	SMPL
O5722-03	EOZ60	980	0.0876 -		SFAM01.1	PPB	12/20/2023 15:20	SMPL
O5722-05	EOZ63	1324	0.1436 -		SFAM01.1	PPB	12/20/2023 15:22	SMPL
O5722-06	EOZ64	709	0.0435 -		SFAM01.1	PPB	12/20/2023 15:24	SMPL
O5722-07	EOZ81	2898	0.3999 -		SFAM01.1	PPB	12/20/2023 15:27	SMPL
O5722-08	EOZ82	2446	0.3263 -		SFAM01.1	PPB	12/20/2023 15:29	SMPL
O5722-09	EOZ83	2041	0.2604 -		SFAM01.1	PPB	12/20/2023 15:31	SMPL
O5722-10	EOZ84	743	0.049 -		SFAM01.1	PPB	12/20/2023 15:33	SMPL
O5722-11	EOZ85	1026	0.0951 -		SFAM01.1	PPB	12/20/2023 15:36	SMPL
O5722-13	EOZ15	768	0.0531 -		SFAM01.1	PPB	12/20/2023 15:38	SMPL
O5722-14	EOZ16	2131	0.275 -		SFAM01.1	PPB	12/20/2023 15:40	SMPL
O5722-15	EOZ17	1551	0.1806 -		SFAM01.1	PPB	12/20/2023 15:43	SMPL

LB128857

INSTRUMENT ID : CV1

O5722-16	E0Z18	1382	0.1531 -	SFAM01.1	PPB	12/20/2023 15:45 SMPL
O5722-17	E0Z19	1192	0.1221 -	SFAM01.1	PPB	12/20/2023 15:47 SMPL
O5722-18	E0Z20	2514	0.3374 -	SFAM01.1	PPB	12/20/2023 15:49 SMPL
O5722-20	E0ZQ1	1200	0.1234 -	SFAM01.1	PPB	12/20/2023 15:52 SMPL
O5722-21	E0ZQ1D	1162	0.1172 -	SFAM01.1	PPB	12/20/2023 15:54 SMPL
O5722-22	E0ZQ1S	16079	2.5462 -	SFAM01.1	PPB	12/20/2023 15:56 SMPL
CCV045	CCV045	29590	4.7462 -	SFAM01.1	PPB	12/20/2023 15:58 SMPL
CCB045	CCB045	49	-0.064 -	SFAM01.1	PPB	12/20/2023 16:01 SMPL
PB157934BL	PBS934	24	-0.0681 -	SFAM01.1	PPB	12/20/2023 16:03 SMPL
O5754-01	E0XM3	622	0.0293 -	SFAM01.1	PPB	12/20/2023 16:05 SMPL
O5754-02	E0XM4	299	-0.0233 -	SFAM01.1	PPB	12/20/2023 16:08 SMPL
O5754-03	E0XS9	915	0.077 -	SFAM01.1	PPB	12/20/2023 16:10 SMPL
O5754-04	E0XT0	19677	3.132 -	SFAM01.1	PPB	12/20/2023 16:12 SMPL
O5754-05	E0XL7	1572	0.184 -	SFAM01.1	PPB	12/20/2023 16:15 SMPL
O5754-06	E0XL9	1204	0.1241 -	SFAM01.1	PPB	12/20/2023 16:17 SMPL
O5754-07	E0XM0	125	-0.0516 -	SFAM01.1	PPB	12/20/2023 16:19 SMPL
O5754-08	E0XT1	590	0.0241 -	SFAM01.1	PPB	12/20/2023 16:21 SMPL
O5754-09	E0XT1D	281	-0.0262 -	SFAM01.1	PPB	12/20/2023 16:24 SMPL
O5754-10	E0XT1S	15483	2.4491 -	SFAM01.1	PPB	12/20/2023 16:26 SMPL
O5754-12	E0XC9	277	-0.0269 -	SFAM01.1	PPB	12/20/2023 16:28 SMPL
O5754-13	E0XL3	547	0.0171 -	SFAM01.1	PPB	12/20/2023 16:31 SMPL
O5754-14	E0XL4	1757	0.2141 -	SFAM01.1	PPB	12/20/2023 16:33 SMPL
O5754-15	E0XL6	115	-0.0532 -	SFAM01.1	PPB	12/20/2023 16:35 SMPL
O5754-16	E0XM5	1912	0.2394 -	SFAM01.1	PPB	12/20/2023 16:37 SMPL
O5754-17	E0XT5	2639	0.3577 -	SFAM01.1	PPB	12/20/2023 16:40 SMPL
O5754-18	E0XD2	145	-0.0484 -	SFAM01.1	PPB	12/20/2023 16:42 SMPL
O5754-19	E0XD3	3913	0.5652 -	SFAM01.1	PPB	12/20/2023 16:44 SMPL
O5754-20	E0XD4	1961	0.2473 -	SFAM01.1	PPB	12/20/2023 16:47 SMPL
CCV046	CCV046	30030	4.8178 -	SFAM01.1	PPB	12/20/2023 16:49 SMPL
CCB046	CCB046	94	-0.0567 -	SFAM01.1	PPB	12/20/2023 16:51 SMPL
O5754-21	E0XM6	585	0.0233 -	SFAM01.1	PPB	12/20/2023 16:53 SMPL
O5754-22	E0XT6	355	-0.0142 -	SFAM01.1	PPB	12/20/2023 16:56 SMPL
CCV047	CCV047	29982	4.81 -	SFAM01.1	PPB	12/20/2023 16:58 SMPL
CCB047	CCB047	69	-0.0607 -	SFAM01.1	PPB	12/20/2023 17:00 SMPL