



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128870

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/21/2023	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	12/21/2023	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	12/21/2023	08:14
4	ICV	1	Water	NA	NA	20.3	7.00	12/21/2023	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	12/21/2023	08:19
6	O5963-01	1	Water	NA	NA	22.2	4.60	12/21/2023	08:25
7	O5963-01DUP	1	Water	NA	NA	22.3	4.61	12/21/2023	08:26
8	O5963-02	1	Water	NA	NA	21.9	4.92	12/21/2023	08:34
9	O5963-03	1	Water	NA	NA	22.4	4.71	12/21/2023	08:35
10	CCV2	1	Water	NA	NA	20.3	12.02	12/21/2023	08:40

WB 124870

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 05812 WATER WorkList ID : 176606 Department : Wet-Chemistry Date : 12-21-2023 07:40:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5963-01	CELL-A	Water	pH	Cool 4 deg C	STER07	O11	12/13/2023	9040C
O5963-02	CELL-B/C	Water	pH	Cool 4 deg C	STER07	O11	12/13/2023	9040C
O5963-03	CELL-D	Water	pH	Cool 4 deg C	STER07	O11	12/13/2023	9040C

12/21/23 08:00

Date/Time
Raw Sample Received by: AB (WC)
Raw Sample Relinquished by: AB sm

Date/Time 12/21/23 10:00
Raw Sample Received by: AB sm
Raw Sample Relinquished by: AB WC