



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 12/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:05
In Date: 12/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:44
Out Date: 12/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB128872

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5893-01	E0XD7	1	1.15	8.76	9.91	8.65	85.6	
O5893-02	E0XD8	2	1.15	8.61	9.76	6.06	57.0	
O5893-03	E0XF9	3	1.13	8.47	9.6	8.92	92.0	
O5893-04	E0XG0	4	1.14	8.55	9.69	8.96	91.5	
O5893-05	E0XG2	5	1.15	8.68	9.83	6.82	65.3	
O5893-06	E0XX0	6	1.15	8.77	9.92	7.17	68.6	
O5893-07	E0XX1	7	1.14	8.41	9.55	5.91	56.7	
O5893-08	E0XH1	8	1.14	8.85	9.99	9.26	91.8	
O5893-09	E0XH2	9	1.14	8.57	9.71	8.96	91.2	
O5893-10	E0XH3	10	1.15	8.81	9.96	8.16	79.6	
O5893-11	E0XH3D	11	1.15	8.81	9.96	8.16	79.6	
O5893-12	E0XH3S	12	1.15	8.81	9.96	8.16	79.6	
O5893-13	E0XH9	13	1.15	8.71	9.86	9.3	93.6	
O5893-14	E0XJ0	14	1.15	8.58	9.73	6.09	57.6	
O5893-15	E0XJ3	15	1.14	8.42	9.56	8.91	92.3	
O5893-16	E0X99	16	1.14	8.73	9.87	9.14	91.6	
O5893-17	E0XA0	17	1.14	8.73	9.87	9.12	91.4	
O5893-18	E0XA1	18	1.15	8.84	9.99	8.99	88.7	
O5893-19	E0XA2	19	1.14	8.49	9.63	7.8	78.4	
O5893-20	E0XJ1	20	1.17	8.82	9.99	9.14	90.4	
O5893-21	E0XJ2	21	1.15	8.44	9.59	7.64	76.9	
O5893-22	E0X53	22	1.14	8.59	9.73	8.93	90.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$