



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128877

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.4	12/21/2023	10:10
2	O5896-05		1	12	100.00	>212.0	>212.0	12/21/2023	10:40
3	O5975-05		1	14	100.00	>212.0	>212.0	12/21/2023	11:10
4	O5975-05DUP		1	14	100.00	>212.0	>212.0	12/21/2023	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66128877

WorkList Name : fp-12-20 WorkList ID : 176574 Department : Wet-Chemistry Date : 12-20-2023 09:36:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5896-05	1104	Water	Flash Point	Cool 4 deg C	PSEG03	L21	12/18/2023	1010B
O5975-05	DS01-GW-122023	Water	Flash Point	Cool 4 deg C	LANG01	O11	12/20/2023	1010B

Date/Time 12/21/2023 10:00
 Raw Sample Received by: RA CWC
 Raw Sample Relinquished by: SP(CWC)

Date/Time 12/21/2023 12:20
 Raw Sample Received by: SP(CWC)
 Raw Sample Relinquished by: RA CWC