

LB128882


SFAM01.1

INSTRUMENT 105. CV2

Linear

 μ Abs.:

52719



A= 0.0000e+000

B= 1.8993e-004 slope

C= -3.8116e-003 y-intercept

Rho= 0.9999746

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/dl
0.00	0.000	0.013	0.013	89	0.000	89					11
0.05	0.050					442	0				15
0.20	0.200	0.230	0.030	1232	0.0 %	1232					-2
2.50	2.500	2.453	-0.047	12933	0.0 %	12933					0
5.00	5.000	4.978	-0.022	26231	0.0 %	26231					0
7.50	7.500	7.516	0.016	39594	0.0 %	39594					0
10.0	10.000	10.009	0.009	52719	0.0 %	52719					0

LB128882

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	89	-		0 SFAM01.1	PPB	12/21/2023 16:45	Std
	0.2 S01	1232	-		0.2 SFAM01.1	PPB	12/21/2023 16:47	Std
	2.5 S02	12933	-		2.5 SFAM01.1	PPB	12/21/2023 16:52	Std
	5 S03	26231	-		5 SFAM01.1	PPB	12/21/2023 16:54	Std
	7.5 S04	39594	-		7.5 SFAM01.1	PPB	12/21/2023 16:56	Std
	10 S05	52719	-		10 SFAM01.1	PPB	12/21/2023 16:59	Std
ICV036	ICV036	20219	3.8365 -		SFAM01.1	PPB	12/21/2023 17:01	SMPL
ICB036	ICB036	-337	-0.0678 -		SFAM01.1	PPB	12/21/2023 17:04	SMPL
CCV051	CCV051	26650	5.0579 -		SFAM01.1	PPB	12/21/2023 17:06	SMPL
CCB051	CCB051	-349	-0.0701 -		SFAM01.1	PPB	12/21/2023 17:08	SMPL
PB157956BL	PBW956	8	-0.0023 -		SFAM01.1	PPB	12/21/2023 17:11	SMPL
PB157986TB	LEB986	57	0.007 -		SFAM01.1	PPB	12/21/2023 17:13	SMPL
O5841-01	E0XD0	104	0.0159 -		SFAM01.1	PPB	12/21/2023 17:15	SMPL
O5841-02	E0XD0D	119	0.0188 -		SFAM01.1	PPB	12/21/2023 17:17	SMPL
O5841-03	E0XD0S	6290	1.1909 -		SFAM01.1	PPB	12/21/2023 17:20	SMPL
O5841-04	E0XL3	-31	-0.0097 -		SFAM01.1	PPB	12/21/2023 17:22	SMPL
O5841-05	E0XA6	386	0.0695 -		SFAM01.1	PPB	12/21/2023 17:24	SMPL
CCV052	CCV052	26532	5.0355 -		SFAM01.1	PPB	12/21/2023 17:27	SMPL
CCB052	CCB052	-241	-0.0496 -		SFAM01.1	PPB	12/21/2023 17:29	SMPL