



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128884

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/22/2023	08:00
2	CAL2	1	Water	NA	NA	20.6	6.98	12/22/2023	08:02
3	CAL3	1	Water	NA	NA	20.3	10.02	12/22/2023	08:05
4	ICV	1	Water	NA	NA	20.5	7.01	12/22/2023	08:10
5	CCV1	1	Water	NA	NA	20.3	2.01	12/22/2023	08:15
6	O5977-05	1	Water	NA	NA	23.2	5.20	12/22/2023	08:20
7	O5992-01	1	Water	NA	NA	21.2	6.71	12/22/2023	08:25
8	O5992-01DUP	1	Water	NA	NA	21.3	6.71	12/22/2023	08:26
9	CCV2	1	Water	NA	NA	20.3	12.02	12/22/2023	08:30

WORKLIST(Hardcopy Internal Chain)

VB 128884

WorkList Name : PH O5992

WorkList ID : 176651

Department : Wet-Chemistry

Date : 12-22-2023 07:41:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5977-05	34725	Water	pH	Cool 4 deg C	PSEG03	O31	12/21/2023	9040C
O5992-01	FM406	Water	pH	Cool 4 deg C	PSEG03	O21	12/21/2023	9040C

Date/Time 12-22-23 07:30

Raw Sample Received by: JQ (woc)

Raw Sample Relinquished by: JQ (woc)

Date/Time 12-22-23

Raw Sample Received by: JQ (woc)

Raw Sample Relinquished by: JQ (woc)

10:00

JQ (woc)

JQ (woc)