

L3128885


7470A

INSTRUMENT ID: 2V1

Linear

μ Abs.:

55432



A= 0.0000e+000

B= 1.8032e-004

C= 4.3128e-002

Rho= 0.9998916

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/o
0.0	0.000	0.051	0.051	41	0.000	41	0				1
0.2	0.200	0.181	-0.019	766	0.0 %	766					10
2.5	2.500	2.442	-0.058	13306	0.0 %	13306					2
5.0	5.000	5.064	0.064	27844	0.0 %	27844					1
7.5	7.500	7.423	-0.077	40927	0.0 %	40927					1
10.0	10.000	10.039	0.039	55432	0.0 %	55432					0

LB128885 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Dec 2023 10:09:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7470A	PPB	22 Dec 2023 10:20:24	S	Std
0.2 - 1	50.2	766	-	0.2000	7470A	PPB	22 Dec 2023 10:22:41	S	Std
2.5 - 1	52.5	13306	-	2.5000	7470A	PPB	22 Dec 2023 10:24:58	S	Std
5.0 - 1	55	27844	-	5.0000	7470A	PPB	22 Dec 2023 10:27:15	S	Std
7.5 - 1	57.5	40927	-	7.5000	7470A	PPB	22 Dec 2023 10:29:34	S	Std
10.0 - 1	52.5	55432	-	10.0000	7470A	PPB	22 Dec 2023 10:31:52	S	Std
ICV37 - 1	ICV37	21926	3.9966	-	7470A	PPB	22 Dec 2023 10:36:26	U	SMPL
ICB37 - 1	ICB37	-288	-0.0088	-	7470A	PPB	22 Dec 2023 10:38:42	U	SMPL
CCV53 - 1	CCV53	27892	5.0726	-	7470A	PPB	22 Dec 2023 10:40:59	U	SMPL
CCB53 - 1	CCB53	-316	-0.0139	-	7470A	PPB	22 Dec 2023 10:43:16	U	SMPL
CRA - 1	CRA	691	0.1677	-	7470A	PPB	22 Dec 2023 10:45:33	U	SMPL
HighStd - 1	HighStd	55546	10.0592	-	7470A	PPB	22 Dec 2023 10:47:48	U	SMPL
ChkStd - 1	ChkStd	38042	6.9029	-	7470A	PPB	22 Dec 2023 10:50:04	U	SMPL
PB157955BL - 1	PBW	-476	-0.0427	-	7470A	PPB	22 Dec 2023 10:52:24	U	SMPL
PB157955BS - 1	LCSW	23122	4.2125	-	7470A	PPB	22 Dec 2023 10:54:44	U	SMPL
O5914-03 - 1	72-12000	-304	-0.0117	-	7470A	PPB	22 Dec 2023 10:57:00	U	SMPL
O5930-02 - 1	CHRT-20413	-18	0.0399	-	7470A	PPB	22 Dec 2023 10:59:18	U	SMPL
O5931-02 - 1	CHRT-24537	-18	0.0399	-	7470A	PPB	22 Dec 2023 11:01:36	U	SMPL
O5931-02DUP - 1	CHRT-24537DUP	40	0.0503	-	7470A	PPB	22 Dec 2023 11:03:52	U	SMPL
O5931-02MS - 1	CHRT-24537MS	23057	4.2008	-	7470A	PPB	22 Dec 2023 11:06:08	U	SMPL
CCV54 - 1	CCV54	27828	5.0611	-	7470A	PPB	22 Dec 2023 11:08:25	U	SMPL
CCB54 - 1	CCB54	-341	-0.0184	-	7470A	PPB	22 Dec 2023 11:10:43	U	SMPL
O5931-02MSD - 1	CHRT-24537MSD	22837	4.1611	-	7470A	PPB	22 Dec 2023 11:13:02	U	SMPL
PB157875TB - 1	PB157875TB	-334	-0.0171	-	7470A	PPB	22 Dec 2023 11:15:17	U	SMPL
O5930-02LX5 - 1		-60	0.0323	-	7470A	PPB	22 Dec 2023 11:17:34	U	SMPL
O5930-02A - 1		21935	3.9985	-	7470A	PPB	22 Dec 2023 11:19:51	U	SMPL
CCV55 - 1	CCV55	27660	5.0308	-	7470A	PPB	22 Dec 2023 11:22:06	U	SMPL
CCB55 - 1	CCB55	-334	-0.0171	-	7470A	PPB	22 Dec 2023 11:24:25	U	SMPL