



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128892

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	12/22/2023	08:45
2	05977-05		1	12	100.00	>212.0	>212.0	12/22/2023	09:15
3	05977-05DUP		1	13	100.00	>212.0	>212.0	12/22/2023	09:45
4	05992-01		1	14	100.00	>212.0	>212.0	12/22/2023	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66128892

WorkList Name : FP1222

WorkList ID : 176662

Department : Wet-Chemistry

Date : 12-22-2023 08:08:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5977-05	34725	Water	Flash Point	Cool 4 deg C	PSEG03	O31	12/21/2023	1010B
O5992-01	FM406	Water	Flash Point	Cool 4 deg C	PSEG03	O21	12/21/2023	1010B

Date/Time 12/22/2023 08:30
 Raw Sample Received by: PM 41617
 Raw Sample Relinquished by: PM 41617

Date/Time 12/22/2023 10:50
 Raw Sample Received by: PM 41617
 Raw Sample Relinquished by: PM 41617