



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128894

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/22/2023	12:35
2	CAL2	1	Water	NA	NA	20.3	7.00	12/22/2023	12:36
3	CAL3	1	Water	NA	NA	20.3	10.02	12/22/2023	12:37
4	ICV	1	Water	NA	NA	20.4	7.00	12/22/2023	12:39
5	CCV1	1	Water	NA	NA	20.3	2.01	12/22/2023	12:40
6	O5997-01	1	Water	NA	NA	21.1	5.24	12/22/2023	12:45
7	O5997-01DUP	1	Water	NA	NA	21.2	5.25	12/22/2023	12:46
8	CCV2	1	Water	NA	NA	20.3	12.03	12/22/2023	12:50

WB 128894

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5997 WATER WorkList ID : 176683 Department : Wet-Chemistry Date : 12-22-2023 12:20:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5997-01	JS9600	Water	pH	Cool 4 deg C	STER17	O11	12/13/2023	9040C

Date/Time 12/12/23 12:30
Raw Sample Received by: JS (WGC)
Raw Sample Relinquished by: JS SH

Date/Time 12/12/23 14:00
Raw Sample Received by: JS SH
Raw Sample Relinquished by: JS SH