



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128895

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/22/2023	10:35
2	CAL2	1	Water	NA	NA	20.3	6.99	12/22/2023	10:36
3	CAL3	1	Water	NA	NA	20.3	10.03	12/22/2023	10:38
4	ICV	1	Water	NA	NA	20.3	7.02	12/22/2023	10:40
5	CCV1	1	Water	NA	NA	20.3	2.01	12/22/2023	10:45
6	O5994-01	1	Solid	20.02	20	21.7	7.55	12/22/2023	11:10
7	O5994-04	1	Solid	20.03	20	21.5	7.66	12/22/2023	11:15
8	O5994-07	1	Solid	20.04	20	21.8	7.82	12/22/2023	11:20
9	O5994-10	1	Solid	20.03	20	21.9	6.89	12/22/2023	11:25
10	O5994-10DUP	1	Solid	20.04	20	21.8	6.90	12/22/2023	11:37
11	CCV2	1	Water	NA	NA	20.3	12.03	12/22/2023	11:40

✓ 128895

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5994

WorkList ID : 176673

Department : Wet-Chemistry

Date : 12-22-2023 10:15:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5994-01	WC-A-COMP	Solid	pH	Cool 4 deg C	EART12	O11	12/21/2023	9045D
O5994-04	WC-B-COMP	Solid	pH	Cool 4 deg C	EART12	O11	12/21/2023	9045D
O5994-07	WC-C-COMP	Solid	pH	Cool 4 deg C	EART12	O11	12/21/2023	9045D
O5994-10	WC-D-COMP	Solid	pH	Cool 4 deg C	EART12	O11	12/21/2023	9045D

Date/Time 12.22.23 10:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 12.22.23 13:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]