

7471B


L13128903

INSTRUMENT ID: CV2

Linear

μ Abs.:

63608



A= 0.0000e+000

B= 1.5563e-004

C= 6.9846e-002

Rho= 0.9988586

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/011
0.0	0.000	0.095	0.095	159	0.000	159					
0.2	0.200	0.214	0.014	924	0.0 %	924					7
2.5	2.500	2.533	0.033	15827	0.0 %	15827					1
5.0	5.000	4.654	-0.346	29456	0.0 %	29456					7
7.5	7.500	7.736	0.236	49257	0.0 %	49257					3
10.0	10.000	9.969	-0.031	63608	0.0 %	63608					0

LB128903 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Dec 2023 13:56:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	159	-	0.0000	7471B	PPB	22 Dec 2023 13:59:44	S	Std
0.2 - 1	50.2	924	-	0.2000	7471B	PPB	22 Dec 2023 14:02:01	S	Std
2.5 - 1	52.5	15827	-	2.5000	7471B	PPB	22 Dec 2023 14:04:17	S	Std
5.0 - 1	55	29456	-	5.0000	7471B	PPB	22 Dec 2023 14:06:34	S	Std
7.5 - 1	59.5	49257	-	7.5000	7471B	PPB	22 Dec 2023 14:08:54	S	Std
10.0 - 1	510	63608	-	10.0000	7471B	PPB	22 Dec 2023 14:11:20	S	Std
ICV38 - 1	ICV38	24540	3.8890	-	7471B	PPB	22 Dec 2023 14:14:39	U	SMPL
ICB38 - 1	ICB38	-175	0.0426	-	7471B	PPB	22 Dec 2023 14:16:54	U	SMPL
CCV56 - 1	CCV56	29503	4.6614	-	7471B	PPB	22 Dec 2023 14:19:16	U	SMPL
CCB56 - 1	CCB56	-197	0.0392	-	7471B	PPB	22 Dec 2023 14:21:31	U	SMPL
CRA - 1	CRA	857	0.2032	-	7471B	PPB	22 Dec 2023 14:23:54	U	SMPL
HighStd - 1	HighStd	64638	10.1294	-	7471B	PPB	22 Dec 2023 14:26:09	U	SMPL
ChkStd - 1	ChkStd	43164	6.7874	-	7471B	PPB	22 Dec 2023 14:28:24	U	SMPL
PB158012BL - 1	PBS	-366	0.0129	-	7471B	PPB	22 Dec 2023 14:30:54	U	SMPL
PB158012BS - 1	LCSS	24761	3.9234	-	7471B	PPB	22 Dec 2023 14:33:21	U	SMPL
O5914-01 - 1	72-12000	1822	0.3534	-	7471B	PPB	22 Dec 2023 14:40:47	U	SMPL
O5914-01DUP - 1	72-12000DUP	1644	0.3257	-	7471B	PPB	22 Dec 2023 14:43:03	U	SMPL
O5914-01MS - 1	72-12000MS	31124	4.9136	-	7471B	PPB	22 Dec 2023 14:45:20	U	SMPL
O5914-01MSD - 1	72-12000MSD	29493	4.6598	-	7471B	PPB	22 Dec 2023 14:50:32	U	SMPL
O5926-01 - 1	MR-319-0-2	1133	0.2462	-	7471B	PPB	22 Dec 2023 14:52:47	U	SMPL
CCV57 - 1	CCV57	29658	4.6855	-	7471B	PPB	22 Dec 2023 14:55:11	U	SMPL
CCB57 - 1	CCB57	-48	0.0624	-	7471B	PPB	22 Dec 2023 14:57:26	U	SMPL
O5926-03 - 1	MR-320-0-2	4482	0.7674	-	7471B	PPB	22 Dec 2023 14:59:52	U	SMPL
O5926-05 - 1	MR-319-3.7-4.7	272	0.1122	-	7471B	PPB	22 Dec 2023 15:02:08	U	SMPL
O5930-01 - 1	CHRT-20413	1473	0.2991	-	7471B	PPB	22 Dec 2023 15:04:25	U	SMPL
O5931-01 - 1	CHRT-24537	7609	1.2540	-	7471B	PPB	22 Dec 2023 15:06:41	U	SMPL
O5933-01 - 1	S-1-SOUTH-SIDE	5238	0.8850	-	7471B	PPB	22 Dec 2023 15:08:57	U	SMPL
O5933-03 - 1	S-2-NORTH-SIDE	6322	1.0537	-	7471B	PPB	22 Dec 2023 15:11:16	U	SMPL
O5940-01 - 1	PL-01-122023	1458	0.2968	-	7471B	PPB	22 Dec 2023 15:13:34	U	SMPL
O5940-03 - 1	PL-02-122023	2432	0.4483	-	7471B	PPB	22 Dec 2023 15:15:54	U	SMPL
O5940-05 - 1	PL-04-122023	1859	0.3592	-	7471B	PPB	22 Dec 2023 15:18:12	U	SMPL
O5956-01 - 1	MC0LL4	5240	0.8853	-	7471B	PPB	22 Dec 2023 15:20:28	U	SMPL
CCV58 - 1	CCV58	29573	4.6723	-	7471B	PPB	22 Dec 2023 15:22:44	U	SMPL
CCB58 - 1	CCB58	-19	0.0669	-	7471B	PPB	22 Dec 2023 15:25:01	U	SMPL
O5956-02 - 1	MC0LL5	9292	1.5159	-	7471B	PPB	22 Dec 2023 15:27:26	U	SMPL
O5956-03 - 1	MC0LT2	1973	0.3769	-	7471B	PPB	22 Dec 2023 15:29:42	U	SMPL
O5956-04 - 1	MC0LW4	4384	0.7521	-	7471B	PPB	22 Dec 2023 15:31:59	U	SMPL
O5956-05 - 1	MC0LM3	5645	0.9484	-	7471B	PPB	22 Dec 2023 15:34:16	U	SMPL
O5969-01 - 1	ETGI-345	6060	1.0130	-	7471B	PPB	22 Dec 2023 15:36:33	U	SMPL
O5970-01 - 1	OK-02-12202023	549	0.1553	-	7471B	PPB	22 Dec 2023 15:38:52	U	SMPL
O5978-01 - 1	N41547	365	0.1267	-	7471B	PPB	22 Dec 2023 15:41:11	U	SMPL
O5980-01 - 1	MR-315-0-2	1008	0.2267	-	7471B	PPB	22 Dec 2023 15:43:29	U	SMPL
O5914-01LX5 - 1		222	0.1044	-	7471B	PPB	22 Dec 2023 15:45:46	U	SMPL
O5914-01A - 1		29602	4.6768	-	7471B	PPB	22 Dec 2023 15:48:01	U	SMPL
CCV59 - 1	CCV59	28863	4.5618	-	7471B	PPB	22 Dec 2023 15:50:17	U	SMPL
CCB59 - 1	CCB59	-81	0.0572	-	7471B	PPB	22 Dec 2023 15:52:41	U	SMPL