



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 12/26/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:05
In Date: 12/23/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 12/24/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB128909

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5922-01	E0X54	1	1.15	8.56	9.71	8.96	91.2	
O5922-02	E0XJ4	2	1.14	8.61	9.75	8.91	90.2	
O5922-03	E0XJ5	3	1.13	8.76	9.89	9.02	90.1	
O5922-04	E0XX3	4	1.15	8.57	9.72	5.74	53.6	
O5922-05	E0XX3D	5	1.15	8.57	9.72	5.74	53.6	
O5922-06	E0XX3S	6	1.15	8.57	9.72	5.74	53.6	
O5922-07	E0X51	7	1.13	8.85	9.98	9.00	88.9	
O5922-08	E0X83	8	1.12	8.87	9.99	9.44	93.8	
O5922-09	E0XK1	9	1.13	8.81	9.94	9.03	89.7	
O5922-10	E0XK2	10	1.13	8.83	9.96	9.00	89.1	
O5922-11	E0XK3	11	1.13	8.86	9.99	9.12	90.2	
O5922-12	E0XK4	12	1.14	8.66	9.8	8.86	89.1	
O5922-13	E0XX7	13	1.15	8.48	9.63	8.08	81.7	
O5922-14	E0X52	14	1.14	8.59	9.73	8.89	90.2	
O5922-15	E0X54	15	1.13	8.69	9.82	9.07	91.4	
O5922-16	E0X84	16	1.15	8.47	9.62	7.61	76.3	
O5922-17	E0X85	17	1.15	8.83	9.98	8.06	78.3	
O5922-18	E0X87	18	1.15	8.84	9.99	9.26	91.7	
O5922-19	E0X88	19	1.15	8.54	9.69	9.07	92.7	
O5922-20	E0XN0	20	1.15	8.84	9.99	8.36	81.6	
O5922-21	E0X86	21	1.15	8.79	9.94	9.2	91.6	
O5922-22	E0X89	22	1.12	8.54	9.66	8.95	91.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$