



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128910

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/23/2023	07:55
2	CAL2	1	Water	NA	NA	20.3	7.00	12/23/2023	07:59
3	CAL3	1	Water	NA	NA	20.3	10.02	12/23/2023	08:00
4	ICV	1	Water	NA	NA	20.4	7.02	12/23/2023	08:10
5	CCV1	1	Water	NA	NA	20.3	2.02	12/23/2023	08:15
6	O6014-01	1	Water	NA	NA	21.7	6.03	12/23/2023	08:20
7	O6014-01DUP	1	Water	NA	NA	21.8	6.04	12/23/2023	08:21
8	O6014-03	1	Water	NA	NA	22.1	5.58	12/23/2023	08:25
9	O6014-05	1	Water	NA	NA	22.6	5.58	12/23/2023	08:30
10	CCV2	1	Water	NA	NA	20.4	12.02	12/23/2023	08:33

18128910

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 06014 WorkList ID : 176698 Department : Wet-Chemistry Date : 12-23-2023 07:42:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6014-01	A AUD-1304	Water	pH	Cool 4 deg C	PSEG03	O31	12/22/2023	9040C
O6014-03	C AUD-1306	Water	pH	Cool 4 deg C	PSEG03	O31	12/22/2023	9040C
O6014-05	C AUD-1307	Water	pH	Cool 4 deg C	PSEG03	O31	12/22/2023	9040C

Date/Time 12/23/23 07:50
Raw Sample Received by: JO COC
Raw Sample Relinquished by: CR SM

Date/Time 12/23/23 11:00
Raw Sample Received by: CR SM
Raw Sample Relinquished by: JO COC