



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128911

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.02	12/23/2023	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	12/23/2023	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	12/23/2023	08:12
4	ICV	1	Water	NA	NA	20.3	6.98	12/23/2023	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	12/23/2023	08:20
6	O6012-01	1	Solid	20.02	20	21.1	8.29	12/23/2023	08:25
7	O6012-01DUP	1	Solid	20.01	20	21.2	8.30	12/22/2023	08:26
8	CCV2	1	Water	NA	NA	20.3	12.02	12/23/2023	08:33

VB 128911

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o6012 WorkList ID : 176697 Department : Wet-Chemistry Date : 12-23-2023 07:41:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6012-01	ASH-SAMPLE	Solid	pH	Cool 4 deg C	G&ST01	O11	12/21/2023	9045D

Date/Time	12/23/23	12:00
Raw Sample Received by:	CP	SM
Raw Sample Relinquished by:	MD	CoC

Page 1 of 1