



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128916

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/23/2023	10:15
2	CAL2	1	Water	NA	NA	20.3	7.00	12/23/2023	10:16
3	CAL3	1	Water	NA	NA	20.3	10.02	12/23/2023	10:19
4	ICV	1	Water	NA	NA	20.3	7.01	12/23/2023	10:20
5	CCV1	1	Water	NA	NA	20.3	2.01	12/23/2023	10:25
6	O6007-01	1	Solid	20.02	20	21.5	10.20	12/23/2023	10:30
7	O6008-01	1	Solid	20.03	20	21.2	7.56	12/23/2023	10:35
8	O6013-01	1	Solid	20.04	20	20.1	4.25	12/23/2023	10:40
9	O6015-02	1	Solid	20.04	20	20.4	8.73	12/23/2023	10:45
10	O6015-04	1	Solid	20.02	20	20.3	8.57	12/23/2023	10:50
11	O6015-06	1	Solid	20.03	20	21.3	10.49	12/23/2023	11:00
12	O6015-08	1	Solid	20.02	20	24.4	10.25	12/23/2023	11:10
13	O6016-02	1	Solid	20.03	20	21.4	10.59	12/23/2023	11:15
14	O6016-04	1	Solid	20.02	20	20.9	10.36	12/23/2023	11:20
15	O6016-04DUP	1	Solid	20.03	20	20.8	10.37	12/23/2023	11:21
16	CCV2	1	Water	NA	NA	20.3	12.02	12/23/2023	11:25

WB 128916

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 06015 WorkList ID : 176715 Department : Wet-Chemistry Date : 12-23-2023 10:05:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6007-01	E8965C-ERM-001	Solid	Corrosivity	Cool 4 deg C	ENTA05	O11	12/22/2023	9045D
O6008-01	LAW-23-00002	Solid	Corrosivity	Cool 4 deg C	PSEG03	O23	12/22/2023	9045D
O6013-01	SOUTHAMPTON-TRANSFORE	Solid	Corrosivity	Cool 4 deg C	PSEG03	O31	12/22/2023	9045D
O6015-02	MR-317-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/21/2023	9045D
O6015-04	MR-317-0-2D	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/21/2023	9045D
O6015-06	MR-323-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/21/2023	9045D
O6015-08	MR-323-6-8	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/21/2023	9045D
O6016-02	MR-322-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/22/2023	9045D
O6016-04	MR-322-4-6	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/22/2023	9045D

Date/Time 12/23/23 10:10
 Raw Sample Received by: JH Coe
 Raw Sample Relinquished by: JH Coe

Date/Time 12/23/23 13:00
 Raw Sample Received by: JH Coe
 Raw Sample Relinquished by: JH Coe