


LB128922

INSTRUMENT ID: CV1

7470A MB
7471B 12/26/23

μ Abs.:

5555E



Linear

A= 0.0000e+000

B= 1.8355e-004

C= -9.3797e-002

Rho= 0.9985792

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/d
0.0	0.000	-0.052	-0.052	230	0.000	230	0				6
0.2	0.200	0.213	0.013	1669	0.0 %	1669					1
2.5	2.500	2.513	0.013	14201	0.0 %	14201					6
5.0	5.000	5.289	0.289	29327	0.0 %	29327					5
7.5	7.500	7.133	-0.367	39373	0.0 %	39373					1
10.0	10.000	10.104	0.104	55558	0.0 %	55558					

LB128922 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Dec 2023 10:22:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	230	-	0.0000	7470A	PPB	26 Dec 2023 10:39:05	S	Std
0.2 - 1	502	1669	-	0.2000	7470A	PPB	26 Dec 2023 10:41:23	S	Std
2.5 - 1	52.5	14201	-	2.5000	7470A	PPB	26 Dec 2023 10:43:40	S	Std
5.0 - 1	55	29327	-	5.0000	7470A	PPB	26 Dec 2023 10:45:57	S	Std
7.5 - 1	59.5	39373	-	7.5000	7470A	PPB	26 Dec 2023 10:48:15	S	Std
10.0 - 1	510	55558	-	10.0000	7470A	PPB	26 Dec 2023 10:50:33	S	Std
ICV40 - 1	ICV40	21760	3.9003	-	7470A	PPB	26 Dec 2023 10:53:30	U	SMPL
ICB40 - 1	ICB40	57	-0.0833	-	7470A	PPB	26 Dec 2023 10:55:47	U	SMPL
CCV64 - 1	CCV64	29190	5.2640	-	7470A	PPB	26 Dec 2023 10:58:04	U	SMPL
CCB64 - 1	CCB64	5	-0.0929	-	7470A	PPB	26 Dec 2023 11:00:19	U	SMPL
CRA - 1	CRA	1603	0.2004	-	7470A	PPB	26 Dec 2023 11:02:37	U	SMPL
HighStd - 1	HighStd	55683	10.1268	-	7470A	PPB	26 Dec 2023 11:04:52	U	SMPL
ChkStd - 1	ChkStd	37999	6.8809	-	7470A	PPB	26 Dec 2023 11:07:08	U	SMPL
PB158033BS - 1	LCSS	21439	3.8413	-	7470A	PPB	26 Dec 2023 11:11:46	U	SMPL
O5896-01 - 1	1104	2302	0.3287	-	7470A	PPB	26 Dec 2023 11:14:02	U	SMPL
O5896-02 - 1	525	789	0.0510	-	7470A	PPB	26 Dec 2023 11:16:21	U	SMPL
O5896-03 - 1	1218	590	0.0145	-	7470A	PPB	26 Dec 2023 11:18:38	U	SMPL
O5896-04 - 1	1108	585	0.0136	-	7470A	PPB	26 Dec 2023 11:20:55	U	SMPL
O5901-01 - 1	GHL-SS-38a	2094	0.2906	-	7470A	PPB	26 Dec 2023 11:23:12	U	SMPL
CCV65 - 1	CCV65	28485	5.1346	-	7470A	PPB	26 Dec 2023 11:25:28	U	SMPL
CCB65 - 1	CCB65	104	-0.0747	-	7470A	PPB	26 Dec 2023 11:27:48	U	SMPL
O5901-01DUP - 1	GHL-SS-38aDUP	2060	0.2843	-	7470A	PPB	26 Dec 2023 11:33:01	U	SMPL
O5901-01MS - 1	GHL-SS-38aMS	19876	3.5545	-	7470A	PPB	26 Dec 2023 11:35:16	U	SMPL
O5901-01MSD - 1	GHL-SS-38aMSD	19955	3.5690	-	7470A	PPB	26 Dec 2023 11:37:31	U	SMPL
O5901-02 - 1	GHL-SS-64a	2617	0.3866	-	7470A	PPB	26 Dec 2023 11:39:49	U	SMPL
O5901-03 - 1	GHL-SS-65a	3522	0.5527	-	7470A	PPB	26 Dec 2023 11:42:06	U	SMPL
O5901-04 - 1	GHL-SS-66a	3372	0.5251	-	7470A	PPB	26 Dec 2023 11:44:22	U	SMPL
O5901-05 - 1	GHL-SS-69a	16380	2.9128	-	7470A	PPB	26 Dec 2023 11:46:39	U	SMPL
O5901-06 - 1	GHL-SS-70a	3450	0.5395	-	7470A	PPB	26 Dec 2023 11:48:56	U	SMPL
O5901-07 - 1	GHL-SS-70aD	3334	0.5182	-	7470A	PPB	26 Dec 2023 11:51:14	U	SMPL
O5901-08 - 1	GHL-SS-71a	4480	0.7285	-	7470A	PPB	26 Dec 2023 11:53:30	U	SMPL
CCV66 - 1	CCV66	28030	5.0511	-	7470A	PPB	26 Dec 2023 11:55:46	U	SMPL
CCB66 - 1	CCB66	120	-0.0718	-	7470A	PPB	26 Dec 2023 11:58:01	U	SMPL
O5901-09 - 1	GHL-SS-72a	19756	3.5324	-	7470A	PPB	26 Dec 2023 12:00:19	U	SMPL
O5901-10 - 1	GHL-SS-73	3851	0.6131	-	7470A	PPB	26 Dec 2023 12:02:35	U	SMPL
O5901-11 - 1	GHL-SS-74a	3515	0.5514	-	7470A	PPB	26 Dec 2023 12:04:52	U	SMPL
O5901-12 - 1	GHL-SS-67a	3132	0.4811	-	7470A	PPB	26 Dec 2023 12:07:07	U	SMPL
O5901-13 - 1	GHL-SS-68a	3725	0.5899	-	7470A	PPB	26 Dec 2023 12:09:24	U	SMPL
PB158034BL - 1	PBS	1	-0.0936	-	7470A	PPB	26 Dec 2023 12:11:40	U	SMPL
PB158034BS - 1	LCSS	21130	3.7846	-	7470A	PPB	26 Dec 2023 12:13:56	U	SMPL
O5902-01 - 1	GHL-SD-44	2597	0.3829	-	7470A	PPB	26 Dec 2023 12:16:14	U	SMPL
O5902-02 - 1	GHL-SD-45	1628	0.2050	-	7470A	PPB	26 Dec 2023 12:18:33	U	SMPL
O5902-03 - 1	GHL-SD-46	3739	0.5925	-	7470A	PPB	26 Dec 2023 12:20:48	U	SMPL
CCV67 - 1	CCV67	27150	4.8896	-	7470A	PPB	26 Dec 2023 12:23:04	U	SMPL
CCB67 - 1	CCB67	113	-0.0731	-	7470A	PPB	26 Dec 2023 12:25:20	U	SMPL
O5902-04 - 1	GHL-SD-47	2191	0.3084	-	7470A	PPB	26 Dec 2023 12:27:37	U	SMPL
O5902-05 - 1	GHL-SD-48	1691	0.2166	-	7470A	PPB	26 Dec 2023 12:29:53	U	SMPL
O5902-06 - 1	GHL-SD-49	5059	0.8348	-	7470A	PPB	26 Dec 2023 12:32:08	U	SMPL
O5902-07 - 1	GHL-SD-50	5637	0.9409	-	7470A	PPB	26 Dec 2023 12:34:24	U	SMPL
O5902-08 - 1	GHL-SD-51	8416	1.4510	-	7470A	PPB	26 Dec 2023 12:36:40	U	SMPL
O5902-09 - 1	GHL-SD-52	5026	0.8287	-	7470A	PPB	26 Dec 2023 12:38:57	U	SMPL
O5902-10 - 1	GHL-SD-53	2394	0.3456	-	7470A	PPB	26 Dec 2023 12:41:15	U	SMPL
O5902-11 - 1	GHL-SD-54	4479	0.7283	-	7470A	PPB	26 Dec 2023 12:43:32	U	SMPL
O5902-12 - 1	GHL-SD-55	3002	0.4572	-	7470A	PPB	26 Dec 2023 12:45:49	U	SMPL
O5902-13 - 1	GHL-SD-56	4126	0.6635	-	7470A	PPB	26 Dec 2023 12:48:05	U	SMPL
CCV68 - 1	CCV68	26605	4.7896	-	7470A	PPB	26 Dec 2023 12:50:21	U	SMPL
CCB68 - 1	CCB68	101	-0.0753	-	7470A	PPB	26 Dec 2023 12:52:37	U	SMPL
O5902-14 - 1	GHL-SD-57	3607	0.5683	-	7470A	PPB	26 Dec 2023 12:54:54	U	SMPL
O5902-15 - 1	GHL-SD-58	3619	0.5705	-	7470A	PPB	26 Dec 2023 12:57:10	U	SMPL
O5902-16 - 1	GHL-SD-59	8270	1.4242	-	7470A	PPB	26 Dec 2023 12:59:25	U	SMPL
O5902-17 - 1	GHL-SD-60	3934	0.6283	-	7470A	PPB	26 Dec 2023 13:01:42	U	SMPL
O5902-18 - 1	GHL-SD-61	21004	3.7615	-	7470A	PPB	26 Dec 2023 13:03:59	U	SMPL
O5902-19 - 1	GHL-SD-62	14505	2.5686	-	7470A	PPB	26 Dec 2023 13:06:16	U	SMPL
O5902-20 - 1	GHL-SD-63	9608	1.6698	-	7470A	PPB	26 Dec 2023 13:13:42	U	SMPL
O5902-20DUP - 1	GHL-SD-63DUP	9033	1.5642	-	7470A	PPB	26 Dec 2023 13:15:58	U	SMPL
O5902-20MS - 1	GHL-SD-63MS	27054	4.8720	-	7470A	PPB	26 Dec 2023 13:18:17	U	SMPL
O5902-20MSD - 1	GHL-SD-63MSD	24647	4.4302	-	7470A	PPB	26 Dec 2023 13:20:34	U	SMPL
CCV69 - 1	CCV69	27789	5.0069	-	7470A	PPB	26 Dec 2023 13:28:02	U	SMPL
CCB69 - 1	CCB69	-4	-0.0945	-	7470A	PPB	26 Dec 2023 13:30:18	U	SMPL
PB158033BL - 1	PBS	537	0.0048	-	7470A	PPB	26 Dec 2023 13:32:36	U	SMPL
O5901-01LX5 - 1		337	-0.0319	-	7470A	PPB	26 Dec 2023 13:34:53	U	SMPL
O5901-01A - 1		19174	3.4256	-	7470A	PPB	26 Dec 2023 13:43:14	U	SMPL
O5902-20LX5 - 1		2038	0.2803	-	7470A	PPB	26 Dec 2023 13:51:56	U	SMPL
O5902-20A - 1		32930	5.9505	-	7470A	PPB	26 Dec 2023 14:01:47	U	SMPL
CCV70 - 1	CCV70	30175	5.4448	-	7470A	PPB	26 Dec 2023 14:09:13	U	SMPL
CCB70 - 1	CCB70	58	-0.0832	-	7470A	PPB	26 Dec 2023 14:11:30	U	SMPL