



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128924

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/26/2023	11:20
2	CAL2	1	Water	NA	NA	20.3	7.00	12/26/2023	11:21
3	CAL3	1	Water	NA	NA	20.3	10.02	12/26/2023	11:22
4	ICV	1	Water	NA	NA	20.3	7.02	12/26/2023	11:25
5	CCV1	1	Water	NA	NA	20.3	2.01	12/26/2023	11:30
6	O5858-15	1	Solid	20.02	20	22.5	10.71	12/26/2023	11:35
7	O5944-01	1	Solid	20.02	20	26.5	10.84	12/26/2023	11:40
8	O5944-03	1	Solid	20.03	20	26.3	10.84	12/26/2023	11:45
9	O5944-04	1	Solid	20.02	20	25.2	11.25	12/26/2023	11:55
10	O5944-05	1	Solid	20.03	20	25.8	10.05	12/26/2023	12:00
11	O5944-06	1	Solid	20.04	20	26.1	11.01	12/26/2023	12:10
12	O5944-08	1	Solid	20.04	20	25.3	9.23	12/26/2023	12:15
13	O5944-09	1	Solid	20.05	20	25.1	8.89	12/26/2023	12:20
14	O5944-10	1	Solid	20.05	20	25.1	8.89	12/26/2023	12:20
15	O5944-11	1	Solid	20.05	20	25.1	8.89	12/26/2023	12:20
16	CCV2	1	Water	NA	NA	20.3	12.02	12/26/2023	12:25
17	O5944-12	1	Solid	20.02	20	25.6	9.20	12/26/2023	12:30
18	O5944-13	1	Solid	20.03	20	24.5	9.64	12/26/2023	12:35
19	O5944-14	1	Solid	20.04	20	24.3	9.25	12/26/2023	12:40
20	CCV3	1	Water	NA	NA	20.3	2.01	12/26/2023	12:45

WORKLIST(Hardcopy Internal Chain)

VB 128924

WorkList Name : ph 05944

WorkList ID : 176745

Department : Wet-Chemistry

Date : 12-26-2023 11:01:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5858-15	E0XC8	Solid	pH	Cool 4 deg C	USEP01	B63	12/12/2023	9045D
O5944-01	E0X90	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-03	E0XX9	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-04	E0XY0	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-05	E0X55	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-06	E0X56	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-08	E0XY1	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-09	E0X57	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-10	E0X57D	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-11	E0X57S	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-12	E0X58	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-13	E0XY6	Solid	pH	Cool 4 deg C	USEP01	B41	12/19/2023	9045D
O5944-14	E0XY7	Solid	pH	Cool 4 deg C	USEP01	B41	12/20/2023	9045D

Date/Time 12/26/23 11:15

Raw Sample Received by: 20 (WOC)

Raw Sample Relinquished by: [Signature]

Date/Time 12/26/23

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: 20 (WOC)