



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB128926

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/26/2023	11:15
2	CAL2	1	Water	NA	NA	20.3	7.00	12/26/2023	11:16
3	CAL3	1	Water	NA	NA	20.3	10.02	12/26/2023	11:19
4	ICV	1	Water	NA	NA	20.5	7.01	12/26/2023	11:23
5	CCV1	1	Water	NA	NA	20.3	2.01	12/26/2023	11:25
6	O5975-05	1	Water	NA	NA	25.2	7.52	12/26/2023	11:30
7	O5975-05DUP	1	Water	NA	NA	25.3	7.53	12/26/2023	11:31
8	CCV2	1	Water	NA	NA	20.3	12.02	12/26/2023	11:35

12/26/23

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSION 05975

WorkList ID : 176741

Department : Wet-Chemistry

Date : 12-26-2023 10:55:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5975-05	DS01-GW-122023	Water	Corrosivity	Cool 4 deg C	LANG01	O11	12/20/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

12/26/23 11:10

12/26/23

12/26/23

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

12/26/23

12/26/23

12/26/23