



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128942

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/26/2023	15:45
2	CAL2	1	Water	NA	NA	20.3	7.01	12/26/2023	15:46
3	CAL3	1	Water	NA	NA	20.3	10.03	12/26/2023	15:50
4	ICV	1	Water	NA	NA	20.3	7.00	12/26/2023	15:55
5	CCV1	1	Water	NA	NA	20.3	2.01	12/26/2023	16:00
6	O6024-01	1	Water	NA	NA	20.8	6.82	12/26/2023	16:05
7	O6024-03	1	Water	NA	NA	20.7	6.69	12/26/2023	16:15
8	O6025-01	1	Water	NA	NA	20.7	5.35	12/26/2023	16:25
9	O6025-02	1	Water	NA	NA	20.7	4.83	12/26/2023	16:30
10	O6025-02DUP	1	Water	NA	NA	20.6	4.84	12/26/2023	16:31
11	CCV2	1	Water	NA	NA	20.3	12.02	12/26/2023	16:35

WORKLIST(Hardcopy Internal Chain)

VB 226123

WorkList Name : ph 06024

WorkList ID : 176765

Department : Wet-Chemistry

Date : 12-26-2023 15:33:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6024-01	C A2421-FRAC-TANK	Water	pH	Cool 4 deg C	PSEG03	O11	12/26/2023	9040C
O6024-03	A TOTE-COMP	Water	pH	Cool 4 deg C	PSEG03	O11	12/26/2023	9040C
O6025-01	C 50200	Water	pH	Cool 4 deg C	PSEG03	O11	12/26/2023	9040C
O6025-02	C 50230	Water	pH	Cool 4 deg C	PSEG03	O11	12/26/2023	9040C

Date/Time 12/26/23 15:40

Raw Sample Received by: JTC

Raw Sample Relinquished by: JTC

Date/Time 12/26/23

Raw Sample Received by: JTC

Raw Sample Relinquished by: JTC