



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128943

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/26/2023	15:45
2	CAL2	1	Water	NA	NA	20.3	7.01	12/26/2023	15:47
3	CAL3	1	Water	NA	NA	20.3	10.03	12/26/2023	15:50
4	ICV	1	Water	NA	NA	20.3	7.00	12/26/2023	15:55
5	CCV1	1	Water	NA	NA	20.3	2.01	12/26/2023	16:40
6	O5785-14	1	Solid	20.02	20	21.2	6.33	12/26/2023	16:45
7	O6025-03	1	Solid	20.01	20	24.2	5.62	12/26/2023	16:50
8	O6025-03DUP	1	Solid	20.02	20	24.3	5.63	12/26/2023	16:51
9	CCV2	1	Water	NA	NA	20.4	12.02	12/26/2023	17:00

16128943

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-o6025 solids

WorkList ID : 176766

Department : Wet-Chemistry

Date : 12-26-2023 15:34:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5785-14	E0Z93	Solid	pH	Cool 4 deg C	USEP01	B41	12/06/2023	9045D
O6025-03	50124	Solid	pH	Cool 4 deg C	PSEG03	O11	12/26/2023	9045D

Date/Time 12/26/23 15:40
Raw Sample Received by: 70 CWCJ
Raw Sample Relinquished by: JTC (SM)

Date/Time 12/26/23 18:00
Raw Sample Received by: JTC (SM)
Raw Sample Relinquished by: 70 CWCJ