



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128949

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	12/27/2023	10:20
2	O6014-01		1	12	100.00	>212.0	>212.0	12/27/2023	11:10
3	O6014-01DUP		1	12	100.00	>212.0	>212.0	12/27/2023	11:40
4	O6014-03		1	14	100.00	>212.0	>212.0	12/27/2023	12:10
5	O6014-05		1	14	100.00	>212.0	>212.0	12/27/2023	12:40
6	O6024-01		1	12	100.00	>212.0	>212.0	12/27/2023	13:10
7	O6024-03		1	14	100.00	>212.0	>212.0	12/27/2023	13:40
8	O6025-01		1	12	100.00	>212.0	>212.0	12/27/2023	14:10
9	O6025-02		1	14	100.00	>212.0	>212.0	12/27/2023	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB128949

WorkList Name : FP-122723-W

WorkList ID : 176782

Department : Wet-Chemistry

Date : 12-27-2023 09:40:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6014-01	AUD-1304	Water	Flash Point	Cool 4 deg C	PSEG03	O31	12/22/2023	1010B
O6014-03	AUD-1306	Water	Flash Point	Cool 4 deg C	PSEG03	O31	12/22/2023	1010B
O6014-05	AUD-1307	Water	Flash Point	Cool 4 deg C	PSEG03	O31	12/22/2023	1010B
O6024-01	A2421-FRAC-TANK	Water	Flash Point	Cool 4 deg C	PSEG03	O11	12/26/2023	1010B
O6024-03	TOTE-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	O11	12/26/2023	1010B
O6025-01	50200	Water	Flash Point	Cool 4 deg C	PSEG03	O11	12/26/2023	1010B
O6025-02	50230	Water	Flash Point	Cool 4 deg C	PSEG03	O11	12/26/2023	1010B

Date/Time 12/27/23 09:45
 Raw Sample Received by: N.P. (W.C.)
 Raw Sample Relinquished by: JH Goe

Date/Time 12/27/23 (14:45)
 Raw Sample Received by: JH Goe
 Raw Sample Relinquished by: N.P. (W.C.)