



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128953

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	12/27/2023	08:15
2	O5977-07		1	14	100.00	>212.0	>212.0	12/27/2023	08:45
3	O5977-07DUP		1	14	100.00	>212.0	>212.0	12/27/2023	09:15
4	O6013-01		1	12	100.00	>212.0	>212.0	12/27/2023	09:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

L13 12-8953

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-122723

WorkList ID : 176781

Department : Wet-Chemistry

Date : 12-27-2023 08:02:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5977-07	34725	Solid	Flash Point	Cool 4 deg C	PSEG03	O31	12/08/2023	1010B
O6013-01	SOUTHAMPTON-TRANSFORE	Solid	Flash Point	Cool 4 deg C	PSEG03	O31	12/22/2023	1010B

Date/Time

12/27/23 08:05

Raw Sample Received by:

N.P (W.C)

Raw Sample Relinquished by:

N.P (W.C)

Date/Time

12/27/23

Raw Sample Received by:

N.P (W.C)

Raw Sample Relinquished by:

N.P (W.C)