



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: Sohil

Run Number: LB128954

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP104779

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB128954BL		1	22.1	0.84	0.89	12/22/2023	13:10
2	LB128954BS	1413	1	22.6	1352.00	1416.95	12/22/2023	13:15
3	O5997-01		1	21.1	3.10	3.35	12/22/2023	13:20
4	O5997-01DUP		1	21.2	3.12	3.36	12/22/2023	13:25

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

WORKLIST(Hardcopy Internal Chain)

LB128954

WorkList Name : CONDUCTANCE-12222:

WorkList ID : 176680

Department : Wet-Chemistry

Date : 12-22-2023 11:17:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5997-01	JS9600	Water	Conductance	Cool 4 deg C	STER17	O11	12/13/2023	SM2510 B

Date/Time 12/22/23 12:40
 Raw Sample Received by: 12/22/23
 Raw Sample Relinquished by: 12/22/23

Date/Time 12/22/23 13:30
 Raw Sample Received by: 12/22/23
 Raw Sample Relinquished by: 12/22/23