



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128955

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/27/2023	16:05
2	CAL2	1	Water	NA	NA	20.3	7.01	12/27/2023	16:06
3	CAL3	1	Water	NA	NA	20.4	10.02	12/27/2023	16:08
4	ICV	1	Water	NA	NA	20.5	7.01	12/27/2023	16:10
5	CCV1	1	Water	NA	NA	20.3	2.01	12/27/2023	16:15
6	O5977-07	1	Solid	20.01	20	23.2	5.20	12/27/2023	16:25
7	CCV2	1	Water	NA	NA	20.3	12.02	12/27/2023	16:30

VB 128955

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5977

WorkList ID : 176803

Department : Wet-Chemistry

Date : 12-27-2023 15:38:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5977-07	34725	Solid	pH	Cool 4 deg C	PSEG03	O31	12/08/2023	9045D

Date/Time 12/27/23 16:00
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC

Date/Time 12/27/23 18:00
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC